

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040220\
 Data File : VX015494.D
 Acq On : 02 Apr 2020 19:53
 Operator : JC/SP
 Sample : PB127948ZHE#16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB127948ZHE#16

Quant Time: Apr 03 09:29:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1071744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1733425	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1658415	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	909612	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	727775	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
35) Dibromofluoromethane	5.47	113	547942	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
50) Toluene-d8	8.70	98	2075430	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	11.12	95	901413	54.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.74%	
Target Compounds						
16) Acetone	2.42	43	46455	7.153	ug/l	97
20) Methylene Chloride	2.84	84	13901	1.146	ug/l	88
43) Isopropyl Acetate	6.42	43	380967	10.852	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040220\
Data File : VX015494.D
Acq On : 02 Apr 2020 19:53
Operator : JC/SP
Sample : PB127948ZHE#16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB127948ZHE#16

Quant Time: Apr 03 09:29:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

