

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027926.D
 Acq On : 02 Apr 2022 12:26
 Operator : JC/MD
 Sample : VX0402WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0402WBSD01

Quant Time: Apr 04 04:13:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64084	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	107136	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46413	56.132	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 112.260%	
35) Dibromofluoromethane	5.385	113	35579	51.449	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.900%	
50) Toluene-d8	8.647	98	119785	49.231	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.460%	
62) 4-Bromofluorobenzene	11.085	95	50996	49.319	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.640%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	11403	19.637	ug/l	98
3) Chloromethane	1.294	50	8841	17.229	ug/l	99
4) Vinyl Chloride	1.374	62	9512	18.075	ug/l	99
5) Bromomethane	1.605	94	5530	23.304	ug/l	94
6) Chloroethane	1.685	64	6865	19.351	ug/l	97
7) Trichlorofluoromethane	1.892	101	17713	19.187	ug/l	98
8) Diethyl Ether	2.136	74	6265	18.906	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.331	101	11978	19.961	ug/l	97
10) Methyl Iodide	2.453	142	11054	15.863	ug/l	# 89
11) Tert butyl alcohol	2.959	59	17270	108.733	ug/l	99
12) 1,1-Dichloroethene	2.319	96	10207	18.539	ug/l	93
13) Acrolein	2.233	56	11713	98.464	ug/l	99
14) Allyl chloride	2.666	41	19456	19.248	ug/l	# 89
15) Acrylonitrile	3.062	53	41683	111.768	ug/l	98
16) Acetone	2.380	43	38469	105.124	ug/l	90
17) Carbon Disulfide	2.514	76	16008	14.213	ug/l	98
18) Methyl Acetate	2.703	43	18603	20.622	ug/l	90
19) Methyl tert-butyl Ether	3.111	73	44821	21.367	ug/l	97
20) Methylene Chloride	2.794	84	12827	19.487	ug/l	94
21) trans-1,2-Dichloroethene	3.099	96	10906	18.250	ug/l	97
22) Diisopropyl ether	3.763	45	44421	21.928	ug/l	# 85
23) Vinyl Acetate	3.721	43	192643	110.069	ug/l	# 90
24) 1,1-Dichloroethane	3.611	63	23386	20.520	ug/l	99
25) 2-Butanone	4.556	43	59500	111.743	ug/l	89
26) 2,2-Dichloropropane	4.477	77	18550	20.379	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	14373	19.946	ug/l	94
28) Bromochloromethane	4.897	49	10373	21.161	ug/l	90
29) Tetrahydrofuran	5.007	42	38089	111.360	ug/l	# 85
30) Chloroform	5.099	83	26637	21.025	ug/l	97
31) Cyclohexane	5.471	56	17078	18.684	ug/l	85
32) 1,1,1-Trichloroethane	5.385	97	21965	20.043	ug/l	98
36) 1,1-Dichloropropene	5.690	75	16309	18.725	ug/l	98
37) Ethyl Acetate	4.715	43	22977	21.280	ug/l	94
38) Carbon Tetrachloride	5.678	117	18862	19.601	ug/l	97
39) Methylcyclohexane	7.385	83	17678	17.248	ug/l	91
40) Benzene	6.044	78	49711	18.833	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027926.D
 Acq On : 02 Apr 2022 12:26
 Operator : JC/MD
 Sample : VX0402WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0402WBSD01

Quant Time: Apr 04 04:13:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	12386	21.566	ug/l	89
42) 1,2-Dichloroethane	6.086	62	22188	21.037	ug/l	93
43) Isopropyl Acetate	6.342	43	33860	19.578	ug/l #	93
44) Trichloroethene	7.123	130	13857	19.318	ug/l	97
45) 1,2-Dichloropropane	7.434	63	13611	19.195	ug/l	100
46) Dibromomethane	7.586	93	10240	19.188	ug/l	91
47) Bromodichloromethane	7.824	83	20307	19.105	ug/l	92
48) Methyl methacrylate	7.696	41	17338	19.116	ug/l #	83
49) 1,4-Dioxane	7.665	88	6994	390.630	ug/l #	89
51) 4-Methyl-2-Pentanone	8.574	43	113845	94.891	ug/l	90
52) Toluene	8.720	92	31771	19.109	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	18435	17.306	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	20507	18.282	ug/l #	83
55) 1,1,2-Trichloroethane	9.153	97	15047	20.405	ug/l	95
56) Ethyl methacrylate	9.116	69	22198	19.007	ug/l #	83
57) 1,3-Dichloropropane	9.311	76	23375	18.836	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	54612	93.660	ug/l	97
59) 2-Hexanone	9.433	43	84926	91.933	ug/l	87
60) Dibromochloromethane	9.525	129	14471	18.926	ug/l	96
61) 1,2-Dibromoethane	9.610	107	14599	18.947	ug/l	98
64) Tetrachloroethene	9.275	164	13080	19.605	ug/l	96
65) Chlorobenzene	10.079	112	36307	18.737	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	13578	18.674	ug/l	99
67) Ethyl Benzene	10.195	91	62502	17.775	ug/l	99
68) m/p-Xylenes	10.305	106	48024	37.134	ug/l	93
69) o-Xylene	10.646	106	24266	19.210	ug/l	93
70) Styrene	10.659	104	40548	18.886	ug/l	95
71) Bromoform	10.799	173	9945	18.617	ug/l #	99
73) Isopropylbenzene	10.963	105	65139	18.130	ug/l	97
74) N-amyl acetate	10.841	43	27547	17.021	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	11.213	83	23468	19.235	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	25819	22.808	ug/l	87
77) Bromobenzene	11.201	156	15556	18.991	ug/l	89
78) n-propylbenzene	11.305	91	75853	17.671	ug/l	98
79) 2-Chlorotoluene	11.366	91	47760	17.705	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	55934	17.941	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.018	75	5693	17.559	ug/l #	70
82) 4-Chlorotoluene	11.457	91	55241	17.845	ug/l	96
83) tert-Butylbenzene	11.719	119	54331	18.022	ug/l	88
84) 1,2,4-Trimethylbenzene	11.756	105	56531	18.185	ug/l	96
85) sec-Butylbenzene	11.890	105	68008	18.280	ug/l	99
86) p-Isopropyltoluene	12.012	119	56025	18.104	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	29178	18.095	ug/l	96
88) 1,4-Dichlorobenzene	12.042	146	29810	17.946	ug/l	97
89) n-Butylbenzene	12.335	91	49366	17.741	ug/l	97
90) Hexachloroethane	12.542	117	8417	16.936	ug/l	90
91) 1,2-Dichlorobenzene	12.335	146	30073	18.881	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.945	75	5663	19.281	ug/l	76
93) 1,2,4-Trichlorobenzene	13.591	180	18151	18.635	ug/l	97
94) Hexachlorobutadiene	13.725	225	8452	20.132	ug/l	93
95) Naphthalene	13.780	128	66700	19.612	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	18049	18.915	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
Data File : VX027926.D
Acq On : 02 Apr 2022 12:26
Operator : JC/MD
Sample : VX0402WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01

Quant Time: Apr 04 04:13:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027926.D
 Acq On : 02 Apr 2022 12:26
 Operator : JC/MD
 Sample : VX0402WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0402WBSD01

Quant Time: Apr 04 04:13:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01

Tgt Ion:168 Resp: 64084
Ion Ratio Lower Upper
168 100
99 67.1 41.7 62.5#






#2
Dichlorodifluoromethane
Concen: 19.637 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 85 Resp: 11403
Ion Ratio Lower Upper
85 100
87 33.6 16.2 48.6



Abundance Scan 34 (1.294 min): VX027779.D\data.ms (-29)



#3

Chloromethane

Concen: 17.229 ug/l

RT: 1.294 min Scan# 34

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :


VX0402WBSD01

Tgt Ion: 50 Resp: 8841




Ion Ratio Lower Upper

50 100

52 33.7 26.4 39.6



Abundance Scan 47 (1.374 min): VX027779.D\data.ms (-42)



#4

Vinyl Chloride

Concen: 18.075 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX027926.D


Acq: 02 Apr 2022 12:26


Tgt Ion: 62 Resp: 9512

Ion Ratio Lower Upper

62 100

64 31.3 25.5 38.3









#5
Bromomethane
Concen: 23.304 ug/l
RT: 1.605 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 94 Resp: 5530
Ion Ratio Lower Upper
94 100
96 99.8 75.3 112.9



#6
Chloroethane
Concen: 19.351 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.007 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 64 Resp: 6865
Ion Ratio Lower Upper
64 100
66 33.1 25.3 37.9









#7
Trichlorofluoromethane
Concen: 19.187 ug/l
RT: 1.892 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:101 Resp: 17713
Ion Ratio Lower Upper
101 100
103 64.3 50.5 75.7



#8
Diethyl Ether
Concen: 18.906 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 74 Resp: 6265
Ion Ratio Lower Upper
74 100
45 88.8 42.0 126.0









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.961 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:	101	Resp:	11978
Ion	Ratio	Lower	Upper
101	100		
85	47.2	33.5	50.3
151	76.7	61.8	92.6



#10
Methyl Iodide
Concen: 15.863 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:	142	Resp:	11054
Ion	Ratio	Lower	Upper
142	100		
127	51.0	33.9	50.9#
141	13.4	11.4	17.2





#11

Tert butyl alcohol

Concen: 108.733 ug/l

RT: 2.959 min Scan# 308

Delta R.T. -0.006 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0402WBSD01

Tgt Ion: 59 Resp: 17270

Ion Ratio Lower Upper

59 100

57 10.8 8.3 12.5



#12

1,1-Dichloroethene

Concen: 18.539 ug/l

RT: 2.319 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26




Tgt Ion: 96 Resp: 10207

Ion Ratio Lower Upper


96 100

61 164.3 124.6 187.0

98 58.2 51.0 76.4




Abundance Scan 188 (2.233 min): VX027779.D\data.ms (-18




m/z-->

Abundance Scan 188 (2.233 min): VX027926.D\data.ms



m/z-->

Abundance Scan 188 (2.233 min): VX027926.D\data.ms (-13



m/z-->

#13

Acrolein

Concen: 98.464 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0402WBSD01


Tgt Ion: 56 Resp: 11713

Ion Ratio Lower Upper

56 100


55 69.5 55.0 82.6

Abundance




Time-->

Abundance Scan 259 (2.666 min): VX027779.D\data.ms (-24




m/z-->

Abundance Scan 259 (2.666 min): VX027926.D\data.ms



m/z-->

Abundance Scan 259 (2.666 min): VX027926.D\data.ms (-21



m/z-->

#14

Allyl chloride

Concen: 19.248 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Tgt Ion: 41 Resp: 19456


Ion Ratio Lower Upper

41 100


39 68.8 50.4 75.6

76 30.5 32.7 49.1#

Abundance







Time-->



#15
Acrylonitrile
Concen: 111.768 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 53 Resp: 41683
Ion Ratio Lower Upper
53 100
52 83.1 65.7 98.5
51 37.7 28.4 42.6



#16
Acetone
Concen: 105.124 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 43 Resp: 38469
Ion Ratio Lower Upper
43 100
58 30.4 28.8 43.2









#17
Carbon Disulfide
Concen: 14.213 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 76 Resp: 16008
Ion Ratio Lower Upper
76 100
78 9.7 7.2 10.8



#18
Methyl Acetate
Concen: 20.622 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 43 Resp: 18603
Ion Ratio Lower Upper
43 100
74 23.3 22.8 34.2









#19
Methyl tert-butyl Ether
Concen: 21.367 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 73 Resp: 44821
Ion Ratio Lower Upper
73 100
57 22.4 16.8 25.2



#20
Methylene Chloride
Concen: 19.487 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 84 Resp: 12827
Ion Ratio Lower Upper
84 100
49 114.1 85.5 128.3
51 29.8 26.6 39.8
86 60.3 51.5 77.3









#21
trans-1,2-Dichloroethene
Concen: 18.250 ug/l
RT: 3.099 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 96 Resp: 10906
Ion Ratio Lower Upper
96 100
61 131.2 105.4 158.2
98 55.2 50.0 75.0



#22
Diisopropyl ether
Concen: 21.928 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 45 Resp: 44421
Ion Ratio Lower Upper
45 100
43 78.6 50.8 76.2#
87 26.2 26.6 39.8#
59 11.3 10.5 15.7









#23
 Vinyl Acetate
 Concen: 110.069 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX027926.D
 Acq: 02 Apr 2022 12:26




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0402WBSD01


Tgt Ion: 43 Resp: 192643
 Ion Ratio Lower Upper
 43 100
 86 9.4 10.8 16.2#



#24
 1,1-Dichloroethane
 Concen: 20.520 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX027926.D
 Acq: 02 Apr 2022 12:26

Tgt Ion: 63 Resp: 23386
 Ion Ratio Lower Upper
 63 100
 98 8.0 3.9 11.7
 100 4.0 2.4 7.2









#25
2-Butanone
Concen: 111.743 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 43 Resp: 59500
Ion Ratio Lower Upper
43 100
72 24.6 24.4 36.6



#26
2,2-Dichloropropane
Concen: 20.379 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 77 Resp: 18550
Ion Ratio Lower Upper
77 100
97 21.3 11.8 35.4









#27
 cis-1,2-Dichloroethene
 Concen: 19.946 ug/l
 RT: 4.495 min Scan# 511
 Delta R.T. 0.000 min
 Lab File: VX027926.D
 Acq: 02 Apr 2022 12:26

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0402WBSD01

Tgt Ion: 96 Resp: 14373




Ion	Ratio	Lower	Upper
96	100		
61	151.6	0.0	283.8
98	65.6	0.0	129.8




#28
 Bromochloromethane
 Concen: 21.161 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. 0.000 min
 Lab File: VX027926.D
 Acq: 02 Apr 2022 12:26

Tgt Ion: 49 Resp: 10373

Ion	Ratio	Lower	Upper
49	100		
129	0.5	0.0	4.4
130	79.1	70.7	106.1





#29

Tetrahydrofuran

Concen: 111.360 ug/l

RT: 5.007 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0402WBSD01





Tgt Ion: 42 Resp: 38089

Ion Ratio Lower Upper

42 100

72 44.1 44.1 66.1#

71 40.8 40.5 60.7



#30

Chloroform

Concen: 21.025 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.000 min

Lab File: VX027926.D




Acq: 02 Apr 2022 12:26


Tgt Ion: 83 Resp: 26637

Ion Ratio Lower Upper

83 100

85 63.1 52.7 79.1









#31
Cyclohexane
Concen: 18.684 ug/l
RT: 5.471 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 56 Resp: 17078
Ion Ratio Lower Upper
56 100
69 29.7 26.7 40.1
84 78.7 77.2 115.8



#32
1,1,1-Trichloroethane
Concen: 20.043 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 97 Resp: 21965
Ion Ratio Lower Upper
97 100
99 64.8 51.8 77.6
61 45.4 34.0 51.0





#33

1,2-Dichloroethane-d4

Concen: 56.132 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0402WBSD01

Tgt Ion: 65 Resp: 46413

Ion Ratio Lower Upper

65 100

67 51.9 0.0 107.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26




Tgt Ion: 114 Resp: 107136


Ion Ratio Lower Upper

114 100

63 22.9 0.0 36.8

88 18.5 0.0 32.4









#35
Dibromofluoromethane
Concen: 51.449 ug/l
RT: 5.385 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:	113	Resp:	35579
Ion	Ratio	Lower	Upper
113	100		
111	103.4	82.8	124.2
192	18.0	16.9	25.3



#36
1,1-Dichloropropene
Concen: 18.725 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:	75	Resp:	16309
Ion	Ratio	Lower	Upper
75	100		
110	33.9	18.3	54.8
77	30.9	24.8	37.2





#37

Ethyl Acetate

Concen: 21.280 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0402WBSD01





Tgt Ion: 43 Resp: 22977

Ion Ratio Lower Upper

43 100

61 13.2 12.6 19.0

70 10.5 10.2 15.2



#38

Carbon Tetrachloride

Concen: 19.601 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26




Tgt Ion: 117 Resp: 18862


Ion Ratio Lower Upper

117 100

119 94.7 77.3 115.9

121 27.3 24.2 36.2





#39

Methylcyclohexane

Concen: 17.248 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0402WBSD01





Tgt Ion: 83 Resp: 17678

Ion Ratio Lower Upper

83 100

55 79.1 54.6 81.8

98 48.1 37.4 56.2



#40

Benzene

Concen: 18.833 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.007 min

Lab File: VX027926.D




Acq: 02 Apr 2022 12:26


Tgt Ion: 78 Resp: 49711

Ion Ratio Lower Upper

78 100

77 21.1 19.3 28.9









#41
Methacrylonitrile
Concen: 21.566 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01

Tgt Ion: 41 Resp: 12386




Ion	Ratio	Lower	Upper
41	100		
39	56.7	43.0	64.4
67	71.5	70.2	105.4
52	31.1	26.0	39.0




#42
1,2-Dichloroethane
Concen: 21.037 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 62 Resp: 22188

Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	20.4









#43
Isopropyl Acetate
Concen: 19.578 ug/l
RT: 6.342 min Scan# 80
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 43 Resp: 33860
Ion Ratio Lower Upper
43 100
61 19.9 18.5 27.7
87 13.0 13.0 19.6#



#44
Trichloroethene
Concen: 19.318 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:130 Resp: 13857
Ion Ratio Lower Upper
130 100
95 97.7 0.0 188.8





#45

1,2-Dichloropropane

Concen: 19.195 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0402WBSD01

Tgt Ion: 63 Resp: 13611

Ion Ratio Lower Upper

63 100

65 32.1 25.8 38.6



#46

Dibromomethane

Concen: 19.188 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26




Tgt Ion: 93 Resp: 10240


Ion Ratio Lower Upper

93 100

95 82.9 65.8 98.6

174 90.5 85.4 128.2









#47
Bromodichloromethane
Concen: 19.105 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:	83	Resp:	20307
Ion	Ratio	Lower	Upper
83	100		
85	57.8	52.2	78.2
127	8.9	6.8	10.2



#48
Methyl methacrylate
Concen: 19.116 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:	41	Resp:	17338
Ion	Ratio	Lower	Upper
41	100		
69	79.8	82.5	123.7#
39	59.9	43.8	65.8





#49

1,4-Dioxane

Concen: 390.630 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX027926.D


Acq: 02 Apr 2022 12:26

Instrument :




MSVOA_X

ClientSampleId :

VX0402WBSD01



Tgt Ion:	88	Resp:	6994
Ion	Ratio	Lower	Upper
88	100		
43	35.1	21.9	32.9#
58	74.7	53.7	80.5



#50

Toluene-d8

Concen: 49.231 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.006 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Tgt Ion:	98	Resp:	119785
Ion	Ratio	Lower	Upper
98	100		
100	64.8	52.6	78.8









#51
4-Methyl-2-Pentanone
Concen: 94.891 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 43 Resp: 113845
Ion Ratio Lower Upper
43 100
58 38.3 36.1 54.1



#52
Toluene
Concen: 19.109 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 92 Resp: 31771
Ion Ratio Lower Upper
92 100
91 173.8 137.5 206.3





#53

t-1,3-Dichloropropene

Concen: 17.306 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0402WBSD01

Tgt Ion: 75 Resp: 18435

Ion Ratio Lower Upper

75 100

77 31.8 25.4 38.0



#54

cis-1,3-Dichloropropene

Concen: 18.282 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26




Tgt Ion: 75 Resp: 20507


Ion Ratio Lower Upper

75 100

77 30.2 24.8 37.2

39 54.6 29.6 44.4#









#55
1,1,2-Trichloroethane
Concen: 20.405 ug/l
RT: 9.153 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 97	Resp: 15047
Ion Ratio	Lower Upper
97 100	
83 88.2	66.2 99.4
85 57.1	42.9 64.3
99 59.4	49.0 73.6



#56
Ethyl methacrylate
Concen: 19.007 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 69	Resp: 22198
Ion Ratio	Lower Upper
69 100	
41 65.6	43.8 65.6
39 37.8	22.2 33.2#









#57
1,3-Dichloropropane
Concen: 18.836 ug/l
RT: 9.311 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 76 Resp: 23375
Ion Ratio Lower Upper
76 100
78 32.6 25.8 38.8



#58
2-Chloroethyl Vinyl ether
Concen: 93.660 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 63 Resp: 54612
Ion Ratio Lower Upper
63 100
106 28.0 23.9 35.9





#59

2-Hexanone

Concen: 91.933 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0402WBSD01

Tgt Ion: 43 Resp: 84926

Ion Ratio Lower Upper

43 100

58 53.5 31.9 95.7



#60

Dibromochloromethane

Concen: 18.926 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX027926.D




Acq: 02 Apr 2022 12:26


Tgt Ion: 129 Resp: 14471

Ion Ratio Lower Upper

129 100

127 77.4 40.3 120.8





#61

1,2-Dibromoethane

Concen: 18.947 ug/l

RT: 9.610 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0402WBSD01

Tgt Ion:107 Resp: 14599

Ion Ratio Lower Upper

107 100

109 95.2 74.4 111.6



#62

4-Bromofluorobenzene

Concen: 49.319 ug/l

RT: 11.085 min Scan# 1640

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26




Tgt Ion: 95 Resp: 50996


Ion Ratio Lower Upper

95 100

174 69.7 0.0 163.0

176 66.6 0.0 155.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:117 Resp: 98398
Ion Ratio Lower Upper
117 100
82 54.8 41.8 62.8
119 30.2 24.8 37.2



#64
Tetrachloroethene
Concen: 19.605 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:164 Resp: 13080
Ion Ratio Lower Upper
164 100
166 127.3 101.2 151.8
129 101.8 73.0 109.4
131 93.5 74.2 111.2









#65
Chlorobenzene
Concen: 18.737 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:112 Resp: 36307
Ion Ratio Lower Upper
112 100
114 31.6 25.8 38.8



#66
1,1,1,2-Tetrachloroethane
Concen: 18.674 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:131 Resp: 13578
Ion Ratio Lower Upper
131 100
133 95.4 48.3 144.8
119 64.4 32.6 98.0





#67

Ethyl Benzene

Concen: 17.775 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0402WBSD01

Tgt Ion: 91 Resp: 62502

Ion Ratio Lower Upper

91 100

106 30.3 24.6 36.8



#68

m/p-Xylenes

Concen: 37.134 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. 0.000 min

Lab File: VX027926.D




Acq: 02 Apr 2022 12:26


Tgt Ion: 106 Resp: 48024

Ion Ratio Lower Upper

106 100

91 211.6 160.4 240.6









#69
o-Xylene
Concen: 19.210 ug/l
RT: 10.646 min Scan# 1567
Delta R.T. 0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:106 Resp: 24266
Ion Ratio Lower Upper
106 100
91 223.8 106.1 318.1



#70
Styrene
Concen: 18.886 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:104 Resp: 40548
Ion Ratio Lower Upper
104 100
78 55.5 38.6 57.8
103 55.3 43.8 65.8









#71
Bromoform
Concen: 18.617 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:173 Resp: 9945
Ion Ratio Lower Upper
173 100
175 48.6 24.1 72.2
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:152 Resp: 48256
Ion Ratio Lower Upper
152 100
115 73.6 37.5 112.7
150 159.2 0.0 338.2









#73
Isopropylbenzene
Concen: 18.130 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:105 Resp: 65139
Ion Ratio Lower Upper
105 100
120 25.7 13.5 40.5



#74
N-ethyl acetate
Concen: 17.021 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 43 Resp: 27547
Ion Ratio Lower Upper
43 100
70 41.2 42.5 63.7#
55 25.1 23.1 34.7
61 23.7 22.8 34.2









#75
1,1,2,2-Tetrachloroethane
Concen: 19.235 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 83 Resp: 23468
Ion Ratio Lower Upper
83 100
131 11.0 5.1 15.3
85 65.4 32.6 97.7



#76
1,2,3-Trichloropropane
Concen: 22.808 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.006 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 75 Resp: 25819
Ion Ratio Lower Upper
75 100
77 30.7 19.4 58.2









#77
Bromobenzene
Concen: 18.991 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:156 Resp: 15556
Ion Ratio Lower Upper
156 100
77 173.6 78.5 235.3
158 93.4 50.2 150.7



#78
n-propylbenzene
Concen: 17.671 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 91 Resp: 75853
Ion Ratio Lower Upper
91 100
120 22.4 11.6 34.8









#79
2-Chlorotoluene
Concen: 17.705 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 91 Resp: 47760
Ion Ratio Lower Upper
91 100
126 32.1 17.3 51.9



#80
1,3,5-Trimethylbenzene
Concen: 17.941 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 105 Resp: 55934
Ion Ratio Lower Upper
105 100
120 46.2 24.3 72.9









#81
trans-1,4-Dichloro-2-butene
Concen: 17.559 ug/l
RT: 11.018 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 75 Resp: 5693
Ion Ratio Lower Upper
75 100
53 125.6 69.0 103.4#
89 50.4 37.0 55.6



#82
4-Chlorotoluene
Concen: 17.845 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 91 Resp: 55241
Ion Ratio Lower Upper
91 100
126 27.5 14.9 44.5









#83
tert-Butylbenzene
Concen: 18.022 ug/l
RT: 11.719 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:	119	Resp:	54331
Ion	Ratio	Lower	Upper
119	100		
91	65.7	26.9	80.6
134	22.6	11.4	34.2



#84
1,2,4-Trimethylbenzene
Concen: 18.185 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:	105	Resp:	56531
Ion	Ratio	Lower	Upper
105	100		
120	42.7	22.6	67.8









#85
sec-Butylbenzene
Concen: 18.280 ug/l
RT: 11.890 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:105 Resp: 68008
Ion Ratio Lower Upper
105 100
134 19.7 10.2 30.4



#86
p-Isopropyltoluene
Concen: 18.104 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:119 Resp: 56025
Ion Ratio Lower Upper
119 100
134 26.3 13.0 39.0
91 29.0 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 18.095 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0402WBSD01





Tgt Ion:146 Resp: 29178

Ion Ratio Lower Upper

146 100

111 44.9 19.7 59.1

148 63.4 31.6 94.7



#88

1,4-Dichlorobenzene

Concen: 17.946 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX027926.D

Acq: 02 Apr 2022 12:26




Tgt Ion:146 Resp: 29810


Ion Ratio Lower Upper

146 100

111 43.4 19.2 57.6

148 64.8 32.3 96.9









#89
n-Butylbenzene
Concen: 17.741 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion: 91	Resp:	49366
Ion	Ratio	Lower Upper
91	100	
92	53.5	27.8 83.3
134	24.8	13.3 39.8



#90
Hexachloroethane
Concen: 16.936 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 117	Resp:	8417
Ion	Ratio	Lower Upper
117	100	
201	70.2	39.5 118.5









#91
1,2-Dichlorobenzene
Concen: 18.881 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:146 Resp: 30073
Ion Ratio Lower Upper
146 100
111 45.3 21.0 63.0
148 63.5 32.6 97.7



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.281 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion: 75 Resp: 5663
Ion Ratio Lower Upper
75 100
155 70.3 47.0 141.2
157 89.7 57.9 173.6









#93
1,2,4-Trichlorobenzene
Concen: 18.635 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01


Tgt Ion:180 Resp: 18151
Ion Ratio Lower Upper
180 100
182 97.5 47.7 143.1
145 34.0 15.2 45.5



#94
Hexachlorobutadiene
Concen: 20.132 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:225 Resp: 8452
Ion Ratio Lower Upper
225 100
223 60.0 34.2 102.5
227 64.3 33.9 101.7









#95
Naphthalene
Concen: 19.612 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Instrument :
MSVOA_X
ClientSampleId :
VX0402WBSD01

Tgt Ion:	128	Resp:	66700
Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.2	15.4
129	10.4	8.7	13.1



#96
1,2,3-Trichlorobenzene
Concen: 18.915 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX027926.D
Acq: 02 Apr 2022 12:26

Tgt Ion:	180	Resp:	18049
Ion	Ratio	Lower	Upper
180	100		
182	98.9	50.0	149.8
145	37.1	16.2	48.5

