

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040222\  
 Data File : VX027927.D  
 Acq On : 02 Apr 2022 12:49  
 Operator : JC/MD  
 Sample : N2074-11RE  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RW8-MW01D3-20220320RE

Quant Time: Apr 04 04:14:16 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032922W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 30 04:39:27 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.556	168	82810	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	143939	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	131390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57323	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63169	59.121	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	118.240%
35) Dibromofluoromethane	5.385	113	49063	52.808	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.620%
50) Toluene-d8	8.653	98	171833	52.565	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.140%
62) 4-Bromofluorobenzene	11.079	95	68134	49.046	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.100%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1988	4.204	ug/l	# 80
29) Tetrahydrofuran	5.007	42	21327	48.253	ug/l	# 84
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040222\  
 Data File : VX027927.D  
 Acq On : 02 Apr 2022 12:49  
 Operator : JC/MD  
 Sample : N2074-11RE  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RW8-MW01D3-20220320RE

Quant Time: Apr 04 04:14:16 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032922W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 30 04:39:27 2022  
 Response via : Initial Calibration

