

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027928.D
 Acq On : 02 Apr 2022 13:12
 Operator : JC/MD
 Sample : N2172-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1R

Quant Time: Apr 04 04:14:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79742	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	138342	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50269	48.857	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.720%
35) Dibromofluoromethane	5.391	113	47227	52.888	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.780%
50) Toluene-d8	8.653	98	165332	52.623	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.240%
62) 4-Bromofluorobenzene	11.085	95	63792	47.778	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.560%
Target Compounds						
16) Acetone	2.380	43	733	1.610	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
Data File : VX027928.D
Acq On : 02 Apr 2022 13:12
Operator : JC/MD
Sample : N2172-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1R

Quant Time: Apr 04 04:14:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

