

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027934.D
 Acq On : 02 Apr 2022 15:32
 Operator : JC/MD
 Sample : N2204-07RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANKRE

Quant Time: Apr 04 04:15:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77596	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	137554	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55016	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61752	61.678	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 123.360%	
35) Dibromofluoromethane	5.385	113	47851	53.894	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.780%	
50) Toluene-d8	8.653	98	164148	52.545	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.080%	
62) 4-Bromofluorobenzene	11.079	95	64030	48.231	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.460%	
Target Compounds						
16) Acetone	2.380	43	1118	2.523	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
Data File : VX027934.D
Acq On : 02 Apr 2022 15:32
Operator : JC/MD
Sample : N2204-07RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANKRE

Quant Time: Apr 04 04:15:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

