

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027935.D
 Acq On : 02 Apr 2022 15:55
 Operator : JC/MD
 Sample : N2237-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-0401

Quant Time: Apr 04 04:15:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76101	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	136226	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61186	62.313	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 124.620%	
35) Dibromofluoromethane	5.385	113	46841	53.270	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.540%	
50) Toluene-d8	8.653	98	164118	53.048	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.100%	
62) 4-Bromofluorobenzene	11.085	95	63828	48.548	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.100%	
Target Compounds						
16) Acetone	2.380	43	919	2.115	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
Data File : VX027935.D
Acq On : 02 Apr 2022 15:55
Operator : JC/MD
Sample : N2237-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-0401

Quant Time: Apr 04 04:15:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

