

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027936.D
 Acq On : 02 Apr 2022 16:19
 Operator : JC/MD
 Sample : N2237-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-0401

Quant Time: Apr 04 04:16:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75618	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	135299	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59854	61.346	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	122.700%
35) Dibromofluoromethane	5.391	113	45828	52.475	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.960%
50) Toluene-d8	8.653	98	162948	53.030	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.060%
62) 4-Bromofluorobenzene	11.085	95	63233	48.425	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.840%
Target Compounds						
16) Acetone	2.373	43	1169	2.707	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
Data File : VX027936.D
Acq On : 02 Apr 2022 16:19
Operator : JC/MD
Sample : N2237-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-0401

Quant Time: Apr 04 04:16:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

