

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027940.D
 Acq On : 02 Apr 2022 17:51
 Operator : JC/MD
 Sample : N2249-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6S

Quant Time: Apr 04 04:17:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79118	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	143068	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61219	59.969	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 119.940%	
35) Dibromofluoromethane	5.391	113	47909	51.879	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.760%	
50) Toluene-d8	8.653	98	170751	52.552	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.100%	
62) 4-Bromofluorobenzene	11.079	95	67332	48.764	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.520%	
Target Compounds						
					Qvalue	
40) Benzene	6.044	78	2302	0.653	ug/l	98
52) Toluene	8.720	92	883	0.398	ug/l	92
67) Ethyl Benzene	10.195	91	3075	0.662	ug/l	100
68) m/p-Xylenes	10.305	106	3870	2.267	ug/l	93
69) o-Xylene	10.646	106	950	0.570	ug/l	70
80) 1,3,5-Trimethylbenzene	11.451	105	881	0.233	ug/l	91
84) 1,2,4-Trimethylbenzene	11.756	105	2822	0.749	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
Data File : VX027940.D
Acq On : 02 Apr 2022 17:51
Operator : JC/MD
Sample : N2249-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6S

Quant Time: Apr 04 04:17:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

