

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027942.D
 Acq On : 02 Apr 2022 18:38
 Operator : JC/MD
 Sample : N2249-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7D

Quant Time: Apr 04 04:17:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77889	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	137462	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57356	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61349	61.045	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 122.080%	
35) Dibromofluoromethane	5.391	113	46928	52.889	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.780%	
50) Toluene-d8	8.653	98	167507	53.656	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.320%	
62) 4-Bromofluorobenzene	11.085	95	65715	49.534	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.060%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1131	1.551	ug/l	96
12) 1,1-Dichloroethene	2.325	96	1362	2.035	ug/l	84
27) cis-1,2-Dichloroethene	4.501	96	344	0.393	ug/l	95
30) Chloroform	5.105	83	3433	2.229	ug/l	91
44) Trichloroethene	7.129	130	3284	3.568	ug/l	94
67) Ethyl Benzene	10.195	91	1643	0.357	ug/l	99
68) m/p-Xylenes	10.305	106	2071	1.225	ug/l	94
69) o-Xylene	10.647	106	610	0.369	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	1229	0.333	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
Data File : VX027942.D
Acq On : 02 Apr 2022 18:38
Operator : JC/MD
Sample : N2249-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7D

Quant Time: Apr 04 04:17:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

