

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX040222\
 Data File : VX027945.D
 Acq On : 02 Apr 2022 19:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 04 04:19:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	50.000	48.931	2.1	71	0.00
3 P	Chloromethane	50.000	47.089	5.8	76	0.00
4 C	Vinyl Chloride	50.000	48.389	3.2#	75	0.00
5 T	Bromomethane	50.000	32.100	35.8#	55	0.00
6 T	Chloroethane	50.000	47.342	5.3	80	0.00
7 T	Trichlorofluoromethane	50.000	49.014	2.0	78	0.00
8 T	Diethyl Ether	50.000	55.332	-10.7	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.538	-3.1	83	0.00
10 T	Methyl Iodide	50.000	32.049	35.9#	50	0.00
11 T	Tert butyl alcohol	250.000	272.981	-9.2	85	0.00
12 CM	1,1-Dichloroethene	50.000	48.795	2.4#	83	0.00
13 T	Acrolein	250.000	301.191	-20.5	95	0.00
14 T	Allyl chloride	50.000	51.155	-2.3	84	0.00
15 T	Acrylonitrile	250.000	292.380	-17.0	90	0.00
16 T	Acetone	250.000	269.653	-7.9	87	0.00
17 T	Carbon Disulfide	50.000	39.339	21.3	68	0.00
18 T	Methyl Acetate	50.000	54.145	-8.3	88	0.00
19 T	Methyl tert-butyl Ether	50.000	56.503	-13.0	89	0.00
20 T	Methylene Chloride	50.000	51.064	-2.1	85	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.431	5.1	80	0.00
22 T	Diisopropyl ether	50.000	57.472	-14.9	101	0.00
23 T	Vinyl Acetate	250.000	290.447	-16.2	101	0.00
24 P	1,1-Dichloroethane	50.000	53.828	-7.7	94	0.00
25 T	2-Butanone	250.000	296.122	-18.4	104	0.00
26 T	2,2-Dichloropropane	50.000	47.825	4.3	82	-0.01
27 T	cis-1,2-Dichloroethene	50.000	52.558	-5.1	85	0.00
28 T	Bromochloromethane	50.000	54.966	-9.9	97	0.00
29 T	Tetrahydrofuran	250.000	293.846	-17.5	102	0.00
30 C	Chloroform	50.000	55.854	-11.7#	95	0.00
31 T	Cyclohexane	50.000	49.021	2.0	82	0.00
32 T	1,1,1-Trichloroethane	50.000	54.286	-8.6	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	55.890	-11.8	91	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	90	0.00
35 S	Dibromofluoromethane	50.000	50.875	-1.8	92	0.00
36 T	1,1-Dichloropropene	50.000	49.670	0.7	87	0.00
37 T	Ethyl Acetate	50.000	52.865	-5.7	101	0.00
38 T	Carbon Tetrachloride	50.000	50.696	-1.4	86	0.00
39 T	Methylcyclohexane	50.000	44.370	11.3	78	0.00
40 TM	Benzene	50.000	48.601	2.8	82	0.00
41 T	Methacrylonitrile	50.000	55.861	-11.7	101	0.00
42 TM	1,2-Dichloroethane	50.000	51.846	-3.7	83	0.00
43 T	Isopropyl Acetate	50.000	52.874	-5.7	82	0.00
44 TM	Trichloroethene	50.000	48.116	3.8	85	0.00
45 C	1,2-Dichloropropane	50.000	50.624	-1.2#	85	0.00
46 T	Dibromomethane	50.000	48.860	2.3	83	0.00
47 T	Bromodichloromethane	50.000	48.666	2.7	82	0.00
48 T	Methyl methacrylate	50.000	49.743	0.5	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
 Data File : VX027945.D
 Acq On : 02 Apr 2022 19:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 04 04:19:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	987.516	1.2	80	0.00
50 S	Toluene-d8	50.000	49.543	0.9	88	0.00
51 T	4-Methyl-2-Pentanone	250.000	260.376	-4.2	85	0.00
52 CM	Toluene	50.000	50.027	-0.1#	85	0.00
53 T	t-1,3-Dichloropropene	50.000	47.538	4.9	78	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.195	1.6	81	0.00
55 T	1,1,2-Trichloroethane	50.000	51.974	-3.9	89	0.00
56 T	Ethyl methacrylate	50.000	51.871	-3.7	84	0.00
57 T	1,3-Dichloropropane	50.000	48.509	3.0	82	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	267.268	-6.9	86	0.00
59 T	2-Hexanone	250.000	241.445	3.4	78	0.00
60 T	Dibromochloromethane	50.000	50.325	-0.7	84	0.00
61 T	1,2-Dibromoethane	50.000	49.059	1.9	83	0.00
62 S	4-Bromofluorobenzene	50.000	46.152	7.7	81	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	90	0.00
64 T	Tetrachloroethene	50.000	47.985	4.0	85	0.00
65 PM	Chlorobenzene	50.000	45.875	8.3	82	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.934	4.1	82	0.00
67 C	Ethyl Benzene	50.000	45.659	8.7#	79	0.00
68 T	m/p-Xylenes	100.000	95.119	4.9	82	0.00
69 T	o-Xylene	50.000	47.979	4.0	81	0.00
70 T	Styrene	50.000	47.967	4.1	82	0.00
71 P	Bromoform	50.000	45.973	8.1	84	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	94	0.00
73 T	Isopropylbenzene	50.000	44.200	11.6	81	0.00
74 T	N-ethyl acetate	50.000	38.231	23.5	66	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.034	-0.1	90	0.00
76 T	1,2,3-Trichloropropane	50.000	48.728	2.5	84	0.00
77 T	Bromobenzene	50.000	50.225	-0.5	90	0.00
78 T	n-propylbenzene	50.000	45.924	8.2	84	0.00
79 T	2-Chlorotoluene	50.000	46.233	7.5	86	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.307	5.4	86	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	33.073	33.9#	59	0.00
82 T	4-Chlorotoluene	50.000	46.562	6.9	85	0.00
83 T	tert-Butylbenzene	50.000	47.740	4.5	88	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.705	2.6	88	0.00
85 T	sec-Butylbenzene	50.000	48.744	2.5	88	0.00
86 T	p-Isopropyltoluene	50.000	49.112	1.8	87	0.00
87 T	1,3-Dichlorobenzene	50.000	48.232	3.5	90	0.00
88 T	1,4-Dichlorobenzene	50.000	47.077	5.8	89	0.00
89 T	n-Butylbenzene	50.000	48.260	3.5	85	0.00
90 T	Hexachloroethane	50.000	46.325	7.3	84	0.00
91 T	1,2-Dichlorobenzene	50.000	48.465	3.1	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.943	0.1	87	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.719	2.6	92	0.00
94 T	Hexachlorobutadiene	50.000	50.006	-0.0	92	0.00
95 T	Naphthalene	50.000	51.768	-3.5	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040222\
Data File : VX027945.D
Acq On : 02 Apr 2022 19:48
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 04 04:19:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.909	-1.8	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6