

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040224\
 Data File : VX040856.D
 Acq On : 02 Apr 2024 11:39
 Operator : JC/MD
 Sample : P1908-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PT-VOA-WP

Quant Time: Apr 03 01:00:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	52776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	90098	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	83218	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42819	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	43828	46.283	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.560%
35) Dibromofluoromethane	5.379	113	29790	46.681	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.360%
50) Toluene-d8	8.647	98	107110	46.449	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.900%
62) 4-Bromofluorobenzene	11.079	95	41494	45.059	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.120%
Target Compounds						
						Qvalue
3) Chloromethane	1.300	50	40639	65.761	ug/l	100
5) Bromomethane	1.611	94	22690	57.817	ug/l	99
10) Methyl Iodide	2.453	142	7054	9.237	ug/l	99
12) 1,1-Dichloroethene	2.325	96	15337	25.976	ug/l	96
16) Acetone	2.373	43	64440	175.256	ug/l	97
20) Methylene Chloride	2.788	84	23719	34.809	ug/l	95
21) trans-1,2-Dichloroethene	3.093	96	22644	36.310	ug/l	98
25) 2-Butanone	4.550	43	36746	67.275	ug/l	98
27) cis-1,2-Dichloroethene	4.483	96	33912	47.091	ug/l	89
30) Chloroform	5.086	83	86515	67.590	ug/l	100
40) Benzene	6.037	78	37703	14.981	ug/l	97
42) 1,2-Dichloroethane	6.086	62	128535	117.590	ug/l	98
44) Trichloroethene	7.122	130	17311	27.541	ug/l	87
45) 1,2-Dichloropropane	7.427	63	21952	36.538	ug/l	95
47) Bromodichloromethane	7.817	83	31350	30.599	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	23162	22.598	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	108396	106.700	ug/l	98
55) 1,1,2-Trichloroethane	9.146	97	21100	34.048	ug/l	91
59) 2-Hexanone	9.427	43	101780	111.264	ug/l	95
60) Dibromochloromethane	9.518	129	67805	91.651	ug/l	98
61) 1,2-Dibromoethane	9.610	107	12575	18.808	ug/l	100
64) Tetrachloroethene	9.274	164	54758	99.433	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	56671	93.810	ug/l	99
67) Ethyl Benzene	10.195	91	159992	49.294	ug/l	98
68) m/p-Xylenes	10.299	106	95746	81.077	ug/l	92
69) o-Xylene	10.640	106	114949	103.321	ug/l	92
70) Styrene	10.652	104	218993	113.593	ug/l	97
71) Bromoform	10.799	173	715	1.858	ug/l #	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	82631	80.807	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	101126	110.753	ug/l	90
80) 1,3,5-Trimethylbenzene	11.451	105	151209	59.200	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	286937	111.515	ug/l	100
88) 1,4-Dichlorobenzene	12.042	146	99577	66.180	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	127887	132.743	ug/l	97
95) Naphthalene	13.774	128	136936	44.143	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040224\
Data File : VX040856.D
Acq On : 02 Apr 2024 11:39
Operator : JC/MD
Sample : P1908-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PT-VOA-WP

Quant Time: Apr 03 01:00:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040224\
Data File : VX040856.D
Acq On : 02 Apr 2024 11:39
Operator : JC/MD
Sample : P1908-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PT-VOA-WP

Quant Time: Apr 03 01:00:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

