

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040224\
 Data File : VX040857.D
 Acq On : 02 Apr 2024 12:17
 Operator : JC/MD
 Sample : PB159929TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159929TB

Quant Time: Apr 03 01:00:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	41889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	74694	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	75703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	37851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	34931	46.475	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.960%
35) Dibromofluoromethane	5.379	113	25849	48.859	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.720%
50) Toluene-d8	8.646	98	95881	50.154	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.300%
62) 4-Bromofluorobenzene	11.079	95	39178	51.318	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.640%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	8324	28.522	ug/l	91
20) Methylene Chloride	2.788	84	20272	37.482	ug/l	97
43) Isopropyl Acetate	6.336	43	37611	26.383	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040224\
Data File : VX040857.D
Acq On : 02 Apr 2024 12:17
Operator : JC/MD
Sample : PB159929TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159929TB

Quant Time: Apr 03 01:00:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

