

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040224\
 Data File : VX040859.D
 Acq On : 02 Apr 2024 14:56
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 03 01:01:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	42138	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.757	114	77209	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	76459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	40469	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	36426	48.178	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.360%
35) Dibromofluoromethane	5.379	113	25947	47.446	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.900%
50) Toluene-d8	8.647	98	96277	48.721	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.440%
62) 4-Bromofluorobenzene	11.079	95	39757	50.380	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040224\
Data File : VX040859.D
Acq On : 02 Apr 2024 14:56
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 03 01:01:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

