



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040224\
Data File : VX040850.D
Acq On : 02 Apr 2024 08:59
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Title : SW846 8260
Last Update : Tue Apr 02 05:40:55 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.4	4626	PASS
75	95	30	60	58.8	12138	PASS
95	95	100	100	100.0	20656	PASS
96	95	5	9	6.6	1366	PASS
173	174	0.00	2	1.2	169	PASS
174	95	50	100	71.1	14695	PASS
175	174	5	9	8.1	1196	PASS
176	174	95	101	97.8	14376	PASS
177	176	5	9	6.0	866	PASS