

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045531.D
 Acq On : 01 Apr 2025 19:25
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 02 03:13:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	58322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	111067	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	100641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	39584	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	54625	51.216	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.440%
35) Dibromofluoromethane	5.385	113	40290	51.128	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.260%
50) Toluene-d8	8.646	98	138308	50.285	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.560%
62) 4-Bromofluorobenzene	11.079	95	51304	51.208	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045531.D
Acq On : 01 Apr 2025 19:25
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 02 03:13:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

