

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045539.D
 Acq On : 02 Apr 2025 12:07
 Operator : JC/MD
 Sample : Q1697-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-19B-72-040125

Quant Time: Apr 03 01:49:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	57807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	116089	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	97826	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	43516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	53760	50.854	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.700%	
35) Dibromofluoromethane	5.385	113	39238	47.639	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.280%	
50) Toluene-d8	8.647	98	127937	44.502	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 89.000%	
62) 4-Bromofluorobenzene	11.079	95	50572	48.294	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 96.580%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	7167	8.818	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.319	101	1062783	1496.426	ug/l	98
12) 1,1-Dichloroethene	2.312	96	59590	85.868	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	7468	10.533	ug/l	92
24) 1,1-Dichloroethane	3.605	63	31180	21.256	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	5175672	5991.947	ug/l	93
32) 1,1,1-Trichloroethane	5.379	97	40840	31.977	ug/l	98
40) Benzene	6.037	78	5720	1.684	ug/l #	88
42) 1,2-Dichloroethane	6.092	62	7366	5.251	ug/l	97
44) Trichloroethene	7.153	130	10380152	12858.147	ug/l	92
52) Toluene	8.720	92	2533	1.232	ug/l	93
64) Tetrachloroethene	9.269	164	512834	742.505	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	13702	9.219	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045539.D
Acq On : 02 Apr 2025 12:07
Operator : JC/MD
Sample : Q1697-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-19B-72-040125

Quant Time: Apr 03 01:49:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

