

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045541.D
 Acq On : 02 Apr 2025 12:54
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 03 01:50:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	60623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	115336	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	107722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44177	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56799	51.233	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.460%		
35) Dibromofluoromethane	5.379	113	42052	51.389	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.780%		
50) Toluene-d8	8.647	98	145187	50.832	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.660%		
62) 4-Bromofluorobenzene	11.079	95	51819	49.808	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 99.620%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	1010	1.356	ug/l #	82
27) cis-1,2-Dichloroethene	4.483	96	2647	2.922	ug/l	88
44) Trichloroethene	7.129	130	10757	13.412	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045541.D
Acq On : 02 Apr 2025 12:54
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 03 01:50:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

