

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045542.D
 Acq On : 02 Apr 2025 13:21
 Operator : JC/MD
 Sample : Q1697-01DL 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-19B-72-040125DL

Quant Time: Apr 03 01:51:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	119039	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58149	51.867	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.740%	
35) Dibromofluoromethane	5.379	113	43629	51.657	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.320%	
50) Toluene-d8	8.647	98	148923	50.518	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.040%	
62) 4-Bromofluorobenzene	11.079	95	54652	50.897	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.800%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	5707	7.577	ug/l	94
12) 1,1-Dichloroethene	2.307	96	382	0.519	ug/l #	56
27) cis-1,2-Dichloroethene	4.483	96	31539	34.430	ug/l	89
44) Trichloroethene	7.117	130	71796	86.731	ug/l	91
64) Tetrachloroethene	9.275	164	3112	3.931	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045542.D
Acq On : 02 Apr 2025 13:21
Operator : JC/MD
Sample : Q1697-01DL 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-19B-72-040125DL

Quant Time: Apr 03 01:51:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

