

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045546.D
 Acq On : 02 Apr 2025 14:54
 Operator : JC/MD
 Sample : Q1697-05 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IW-03-55-040125

Quant Time: Apr 03 01:52:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	58143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	113862	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	102695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	39273	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57174	53.771	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.540%	
35) Dibromofluoromethane	5.385	113	41733	51.659	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.320%	
50) Toluene-d8	8.646	98	141159	50.061	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.120%	
62) 4-Bromofluorobenzene	11.079	95	48414	47.137	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 94.280%	
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	707	1.619	ug/l	90
27) cis-1,2-Dichloroethene	4.489	96	2922	3.363	ug/l	90
44) Trichloroethene	7.128	130	6502	8.212	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045546.D
Acq On : 02 Apr 2025 14:54
Operator : JC/MD
Sample : Q1697-05 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IW-03-55-040125

Quant Time: Apr 03 01:52:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

