

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045550.D
 Acq On : 02 Apr 2025 16:27
 Operator : JC/MD
 Sample : Q1697-04 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IW-02-55-040125-FD

Quant Time: Apr 03 01:54:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66382	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130300	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63193	52.056	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.120%	
35) Dibromofluoromethane	5.379	113	46257	50.036	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.080%	
50) Toluene-d8	8.647	98	161403	50.020	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.040%	
62) 4-Bromofluorobenzene	11.079	95	60106	51.138	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.280%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.306	96	594	0.745	ug/l	92
16) Acetone	2.386	43	8926	17.906	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	32054	32.316	ug/l	91
29) Tetrahydrofuran	5.007	42	10809	22.848	ug/l	96
44) Trichloroethene	7.123	130	55551	61.307	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045550.D
Acq On : 02 Apr 2025 16:27
Operator : JC/MD
Sample : Q1697-04 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IW-02-55-040125-FD

Quant Time: Apr 03 01:54:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

