

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045552.D
 Acq On : 02 Apr 2025 17:14
 Operator : JC/MD
 Sample : Q1697-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-040125

Quant Time: Apr 03 01:55:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	59888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	115627	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105702	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	38388	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57754	52.734	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.460%		
35) Dibromofluoromethane	5.385	113	42932	52.332	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.660%		
50) Toluene-d8	8.646	98	147358	51.462	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.920%		
62) 4-Bromofluorobenzene	11.079	95	49475	47.435	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 94.880%		
Target Compounds						
						Qvalue
16) Acetone	2.385	43	702	1.561	ug/l #	88
44) Trichloroethene	7.128	130	821	1.021	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045552.D
Acq On : 02 Apr 2025 17:14
Operator : JC/MD
Sample : Q1697-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-040125

Quant Time: Apr 03 01:55:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

