

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045532.D
 Acq On : 01 Apr 2025 19:48
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX040225

Quant Time: Apr 02 03:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	52.454	-4.9	97	0.00
3 P	Chloromethane	50.000	48.349	3.3	96	0.00
4 C	Vinyl Chloride	50.000	48.816	2.4#	97	0.00
5 T	Bromomethane	50.000	46.991	6.0	96	0.00
6 T	Chloroethane	50.000	51.586	-3.2	98	0.00
7 T	Trichlorofluoromethane	50.000	50.354	-0.7	99	0.00
8 T	Diethyl Ether	50.000	49.228	1.5	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.528	0.9	99	0.00
10 T	Methyl Iodide	50.000	48.626	2.7	95	0.00
11 T	Tert butyl alcohol	250.000	246.612	1.4	97	0.00
12 CM	1,1-Dichloroethene	50.000	49.726	0.5#	100	0.00
13 T	Acrolein	250.000	238.141	4.7	98	0.00
14 T	Allyl chloride	50.000	50.541	-1.1	100	0.00
15 T	Acrylonitrile	250.000	248.351	0.7	96	0.00
16 T	Acetone	250.000	241.904	3.2	98	0.00
17 T	Carbon Disulfide	50.000	52.057	-4.1	101	0.00
18 T	Methyl Acetate	50.000	48.742	2.5	100	0.00
19 T	Methyl tert-butyl Ether	50.000	50.950	-1.9	102	0.00
20 T	Methylene Chloride	50.000	49.055	1.9	99	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.800	0.4	99	0.00
22 T	Diisopropyl ether	50.000	50.895	-1.8	101	0.00
23 T	Vinyl Acetate	250.000	267.449	-7.0	101	0.00
24 P	1,1-Dichloroethane	50.000	49.677	0.6	101	0.00
25 T	2-Butanone	250.000	254.882	-2.0	97	0.00
26 T	2,2-Dichloropropane	50.000	52.020	-4.0	102	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.049	1.9	99	0.00
28 T	Bromochloromethane	50.000	48.403	3.2	101	0.00
29 T	Tetrahydrofuran	250.000	249.473	0.2	97	0.00
30 C	Chloroform	50.000	49.614	0.8#	100	0.00
31 T	Cyclohexane	50.000	49.583	0.8	100	0.00
32 T	1,1,1-Trichloroethane	50.000	50.330	-0.7	102	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.329	-0.7	107	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	104	0.00
35 S	Dibromofluoromethane	50.000	49.256	1.5	105	0.00
36 T	1,1-Dichloropropene	50.000	50.088	-0.2	103	0.00
37 T	Ethyl Acetate	50.000	49.787	0.4	100	0.00
38 T	Carbon Tetrachloride	50.000	50.667	-1.3	102	0.00
39 T	Methylcyclohexane	50.000	50.124	-0.2	98	0.00
40 TM	Benzene	50.000	49.212	1.6	101	0.00
41 T	Methacrylonitrile	50.000	53.098	-6.2	100	0.00
42 TM	1,2-Dichloroethane	50.000	48.875	2.3	99	0.00
43 T	Isopropyl Acetate	50.000	51.539	-3.1	100	0.00
44 TM	Trichloroethene	50.000	47.643	4.7	98	0.00
45 C	1,2-Dichloropropane	50.000	49.521	1.0#	100	0.00
46 T	Dibromomethane	50.000	48.542	2.9	97	0.00
47 T	Bromodichloromethane	50.000	49.273	1.5	100	0.00
48 T	Methyl methacrylate	50.000	51.507	-3.0	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045532.D
 Acq On : 01 Apr 2025 19:48
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX040225

Quant Time: Apr 02 03:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	955.327	4.5	92	0.00
50 S	Toluene-d8	50.000	50.160	-0.3	106	0.00
51 T	4-Methyl-2-Pentanone	250.000	254.994	-2.0	97	0.00
52 CM	Toluene	50.000	48.994	2.0#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	46.783	6.4	103	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.878	-7.8	104	0.00
55 T	1,1,2-Trichloroethane	50.000	49.007	2.0	100	0.00
56 T	Ethyl methacrylate	50.000	53.705	-7.4	102	0.00
57 T	1,3-Dichloropropane	50.000	49.869	0.3	101	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	260.201	-4.1	100	0.00
59 T	2-Hexanone	250.000	251.710	-0.7	96	0.00
60 T	Dibromochloromethane	50.000	50.983	-2.0	99	0.00
61 T	1,2-Dibromoethane	50.000	50.693	-1.4	101	0.00
62 S	4-Bromofluorobenzene	50.000	50.959	-1.9	105	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	101	0.00
64 T	Tetrachloroethene	50.000	47.239	5.5	97	0.00
65 PM	Chlorobenzene	50.000	50.207	-0.4	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.495	-1.0	102	0.00
67 C	Ethyl Benzene	50.000	51.558	-3.1#	100	0.00
68 T	m/p-Xylenes	100.000	102.728	-2.7	100	0.00
69 T	o-Xylene	50.000	51.650	-3.3	100	0.00
70 T	Styrene	50.000	52.624	-5.2	99	0.00
71 P	Bromoform	50.000	52.694	-5.4	100	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	52.794	-5.6	99	0.00
74 T	N-amyl acetate	50.000	54.387	-8.8	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.608	2.8	97	0.00
76 T	1,2,3-Trichloropropane	50.000	49.741	0.5	97	0.00
77 T	Bromobenzene	50.000	50.525	-1.0	98	0.00
78 T	n-propylbenzene	50.000	53.967	-7.9	100	0.00
79 T	2-Chlorotoluene	50.000	51.057	-2.1	100	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.534	-7.1	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.482	-5.0	100	0.00
82 T	4-Chlorotoluene	50.000	52.422	-4.8	98	0.00
83 T	tert-Butylbenzene	50.000	53.274	-6.5	100	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.990	-6.0	98	0.00
85 T	sec-Butylbenzene	50.000	53.903	-7.8	99	0.00
86 T	p-Isopropyltoluene	50.000	54.337	-8.7	101	0.00
87 T	1,3-Dichlorobenzene	50.000	50.521	-1.0	98	0.00
88 T	1,4-Dichlorobenzene	50.000	49.117	1.8	97	0.00
89 T	n-Butylbenzene	50.000	55.189	-10.4	99	0.00
90 T	Hexachloroethane	50.000	54.066	-8.1	102	0.00
91 T	1,2-Dichlorobenzene	50.000	49.938	0.1	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.226	-6.5	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.935	-3.9	101	0.00
94 T	Hexachlorobutadiene	50.000	49.961	0.1	98	0.00
95 T	Naphthalene	50.000	52.765	-5.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.396	-0.8	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6