

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040319\
 Data File : VX008611.D
 Acq On : 03 Apr 2019 11:52
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 04 05:50:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	198894	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	332896	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	292408	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	128030	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	147384	50.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.60%	
35) Dibromofluoromethane	5.50	113	105633	49.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.26%	
50) Toluene-d8	8.71	98	425510	50.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.42%	
62) 4-Bromofluorobenzene	11.13	95	148636	49.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040319\
Data File : VX008611.D
Acq On : 03 Apr 2019 11:52
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 04 05:50:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

