

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034918.D
 Acq On : 03 Apr 2023 09:43
 Operator : JC/MD
 Sample : VX0403MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0403MBL01

Quant Time: Apr 03 21:32:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	285824	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	487223	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	432504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	191991	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	212276	48.562	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.120%
35) Dibromofluoromethane	5.379	113	156945	48.701	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.400%
50) Toluene-d8	8.647	98	582839	49.000	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.000%
62) 4-Bromofluorobenzene	11.079	95	220765	47.961	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
Data File : VX034918.D
Acq On : 03 Apr 2023 09:43
Operator : JC/MD
Sample : VX0403MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0403MBL01

Quant Time: Apr 03 21:32:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

