

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034926.D
 Acq On : 03 Apr 2023 12:51
 Operator : JC/MD
 Sample : 02159-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Apr 03 21:35:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	275137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	467927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	411640	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	204388	48.573	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.140%
35) Dibromofluoromethane	5.385	113	149835	48.413	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.820%
50) Toluene-d8	8.647	98	561217	49.127	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.260%
62) 4-Bromofluorobenzene	11.079	95	201436	45.567	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.140%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1720	0.510	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
Data File : VX034926.D
Acq On : 03 Apr 2023 12:51
Operator : JC/MD
Sample : 02159-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Apr 03 21:35:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

