

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034928.D
 Acq On : 03 Apr 2023 13:38
 Operator : JC/MD
 Sample : 02137-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Apr 03 21:36:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	266383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	454791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	408489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	203444	49.938	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.880%
35) Dibromofluoromethane	5.385	113	142888	47.501	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.000%
50) Toluene-d8	8.647	98	553674	49.867	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.740%
62) 4-Bromofluorobenzene	11.079	95	204110	47.505	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.020%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	9939	6.006	ug/l	92
20) Methylene Chloride	2.794	84	4917	1.506	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
Data File : VX034928.D
Acq On : 03 Apr 2023 13:38
Operator : JC/MD
Sample : 02137-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Apr 03 21:36:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

