

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034929.D
 Acq On : 03 Apr 2023 14:01
 Operator : JC/MD
 Sample : 02163-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB

Quant Time: Apr 03 21:36:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	272318	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	464773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	413840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178782	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	205148	49.259	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.520%
35) Dibromofluoromethane	5.385	113	147736	48.058	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	8.647	98	557139	49.101	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.079	95	204594	46.595	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
Data File : VX034929.D
Acq On : 03 Apr 2023 14:01
Operator : JC/MD
Sample : 02163-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB

Quant Time: Apr 03 21:36:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

