

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034932.D
 Acq On : 03 Apr 2023 15:10
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 03 21:38:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	263701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	453254	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	399439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	175947	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	202258	50.152	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.300%		
35) Dibromofluoromethane	5.385	113	141584	47.228	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 94.460%		
50) Toluene-d8	8.647	98	537807	48.602	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.200%		
62) 4-Bromofluorobenzene	11.079	95	202513	47.293	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 94.580%		
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	2565	0.793	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
Data File : VX034932.D
Acq On : 03 Apr 2023 15:10
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 03 21:38:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

