

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034933.D
 Acq On : 03 Apr 2023 15:33
 Operator : JC/MD
 Sample : 02080-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T-5

Quant Time: Apr 03 21:38:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	265740	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	459884	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	411510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181617	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	203441	50.058	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.120%
35) Dibromofluoromethane	5.385	113	143760	47.262	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.520%
50) Toluene-d8	8.647	98	552732	49.231	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.460%
62) 4-Bromofluorobenzene	11.079	95	207808	47.830	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.660%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	8136	4.928	ug/l	90
19) Methyl tert-butyl Ether	3.111	73	951043	94.157	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
Data File : VX034933.D
Acq On : 03 Apr 2023 15:33
Operator : JC/MD
Sample : 02080-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T-5

Quant Time: Apr 03 21:38:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

