

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034937.D
 Acq On : 03 Apr 2023 17:06
 Operator : JC/MD
 Sample : 02161-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-5

Quant Time: Apr 03 21:40:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	265904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	462338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	421083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	186546	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	201687	49.596	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.200%
35) Dibromofluoromethane	5.391	113	147782	48.326	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.660%
50) Toluene-d8	8.647	98	560853	49.689	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.380%
62) 4-Bromofluorobenzene	11.079	95	212529	48.657	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.320%
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.130	74	352638	189.244	ug/l	94
16) Acetone	2.380	43	109461	66.260	ug/l	95
20) Methylene Chloride	2.788	84	24813	7.612	ug/l	97
25) 2-Butanone	4.562	43	12406	5.163	ug/l	97
37) Ethyl Acetate	4.714	43	78779	16.630	ug/l	98
43) Isopropyl Acetate	6.342	43	271776	34.292	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
Data File : VX034937.D
Acq On : 03 Apr 2023 17:06
Operator : JC/MD
Sample : 02161-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-5

Quant Time: Apr 03 21:40:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

