

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034917.D
 Acq On : 03 Apr 2023 09:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 04 12:47:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.499	0.547	-9.6	121	0.00
3 P	Chloromethane	0.624	0.611	2.1	106	0.00
4 C	Vinyl Chloride	0.618	0.608	1.6#	107	0.00
5 T	Bromomethane	0.402	0.350	12.9	94	0.00
6 T	Chloroethane	0.402	0.342	14.9	93	0.00
7 T	Trichlorofluoromethane	0.994	0.883	11.2	97	0.00
8 T	Diethyl Ether	0.350	0.332	5.1	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.519	0.529	-1.9	110	0.00
10 T	Methyl Iodide	0.647	0.560	13.4	89	0.00
11 T	Tert butyl alcohol	0.134	0.097	27.6#	75	-0.01
12 CM	1,1-Dichloroethene	0.513	0.484	5.7#	102	0.00
13 T	Acrolein	0.138	0.117	15.2	90	0.00
14 T	Allyl chloride	1.006	0.994	1.2	106	0.00
15 T	Acrylonitrile	0.302	0.273	9.6	95	0.00
16 T	Acetone	0.311	0.360	-15.8	124	0.00
17 T	Carbon Disulfide	1.286	1.244	3.3	101	0.00
18 T	Methyl Acetate	1.104	1.021	7.5	99	0.00
19 T	Methyl tert-butyl Ether	1.900	1.825	3.9	103	0.00
20 T	Methylene Chloride	0.613	0.555	9.5	101	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.532	5.0	101	0.00
22 T	Diisopropyl ether	2.029	1.965	3.2	103	0.00
23 T	Vinyl Acetate	1.447	1.434	0.9	100	0.00
24 P	1,1-Dichloroethane	1.092	1.050	3.8	103	0.00
25 T	2-Butanone	0.452	0.444	1.8	101	0.00
26 T	2,2-Dichloropropane	0.898	0.893	0.6	103	0.00
27 T	cis-1,2-Dichloroethene	0.663	0.623	6.0	102	0.00
28 T	Bromochloromethane	0.534	0.491	8.1	97	0.00
29 T	Tetrahydrofuran	0.283	0.252	11.0	93	0.00
30 C	Chloroform	1.109	1.074	3.2#	102	0.00
31 T	Cyclohexane	0.931	0.937	-0.6	104	0.00
32 T	1,1,1-Trichloroethane	0.946	0.962	-1.7	105	0.00
33 S	1,2-Dichloroethane-d4	0.765	0.731	4.4	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.331	0.319	3.6	96	0.00
36 T	1,1-Dichloropropene	0.490	0.492	-0.4	104	0.00
37 T	Ethyl Acetate	0.512	0.481	6.1	93	0.00
38 T	Carbon Tetrachloride	0.462	0.504	-9.1	109	0.00
39 T	Methylcyclohexane	0.520	0.526	-1.2	104	0.00
40 TM	Benzene	1.406	1.355	3.6	100	0.00
41 T	Methacrylonitrile	0.279	0.273	2.2	97	0.00
42 TM	1,2-Dichloroethane	0.559	0.552	1.3	102	0.00
43 T	Isopropyl Acetate	0.857	0.817	4.7	96	0.00
44 TM	Trichloroethene	0.359	0.358	0.3	103	0.00
45 C	1,2-Dichloropropane	0.367	0.359	2.2#	101	0.00
46 T	Dibromomethane	0.252	0.244	3.2	99	0.00
47 T	Bromodichloromethane	0.466	0.495	-6.2	104	0.00
48 T	Methyl methacrylate	0.421	0.415	1.4	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034917.D
 Acq On : 03 Apr 2023 09:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 04 12:47:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	88	-0.01
50 S	Toluene-d8	1.221	1.186	2.9	96	0.00
51 T	4-Methyl-2-Pentanone	0.507	0.480	5.3	95	0.00
52 CM	Toluene	0.917	0.861	6.1#	101	0.00
53 T	t-1,3-Dichloropropene	0.501	0.542	-8.2	101	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.583	-6.4	101	0.00
55 T	1,1,2-Trichloroethane	0.347	0.338	2.6	99	0.00
56 T	Ethyl methacrylate	0.544	0.544	0.0	96	0.00
57 T	1,3-Dichloropropane	0.615	0.602	2.1	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.277	0.279	-0.7	98	0.00
59 T	2-Hexanone	0.384	0.381	0.8	97	0.00
60 T	Dibromochloromethane	0.331	0.364	-10.0	105	0.00
61 T	1,2-Dibromoethane	0.361	0.355	1.7	98	0.00
62 S	4-Bromofluorobenzene	0.472	0.473	-0.2	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.358	0.335	6.4	100	0.00
65 PM	Chlorobenzene	1.040	1.019	2.0	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.358	0.380	-6.1	103	0.00
67 C	Ethyl Benzene	1.885	1.863	1.2#	101	0.00
68 T	m/p-Xylenes	0.720	0.712	1.1	102	0.00
69 T	o-Xylene	0.705	0.692	1.8	101	0.00
70 T	Styrene	1.126	1.164	-3.4	102	0.00
71 P	Bromoform	0.243	0.269	-10.7	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.663	3.689	-0.7	103	0.00
74 T	N-amyl acetate	1.647	1.646	0.1	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.200	1.126	6.2	97	0.00
76 T	1,2,3-Trichloropropane	1.007	1.264	-25.5#	129	0.00
77 T	Bromobenzene	0.908	0.861	5.2	100	0.00
78 T	n-propylbenzene	4.231	4.276	-1.1	104	0.00
79 T	2-Chlorotoluene	2.671	2.625	1.7	101	0.00
80 T	1,3,5-Trimethylbenzene	3.062	3.084	-0.7	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.336	0.332	1.2	95	0.00
82 T	4-Chlorotoluene	3.102	3.067	1.1	102	0.00
83 T	tert-Butylbenzene	2.876	2.943	-2.3	104	0.00
84 T	1,2,4-Trimethylbenzene	3.114	3.106	0.3	102	0.00
85 T	sec-Butylbenzene	3.498	3.579	-2.3	104	0.00
86 T	p-Isopropyltoluene	2.917	3.049	-4.5	105	0.00
87 T	1,3-Dichlorobenzene	1.653	1.650	0.2	102	0.00
88 T	1,4-Dichlorobenzene	1.726	1.675	3.0	101	0.00
89 T	n-Butylbenzene	2.549	2.650	-4.0	104	0.00
90 T	Hexachloroethane	0.452	0.513	-13.5	106	0.00
91 T	1,2-Dichlorobenzene	1.655	1.617	2.3	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.251	0.253	-0.8	93	0.00
93 T	1,2,4-Trichlorobenzene	0.936	0.976	-4.3	104	0.00
94 T	Hexachlorobutadiene	0.412	0.403	2.2	109	0.00
95 T	Naphthalene	3.326	3.301	0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
Data File : VX034917.D
Acq On : 03 Apr 2023 09:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 04 12:47:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.944	0.960	-1.7	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6