

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
 Data File : VX034919.D
 Acq On : 03 Apr 2023 10:06
 Operator : JC/MD
 Sample : VX0403WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0403WBL01

Quant Time: Apr 03 21:32:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	297985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	504919	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	449098	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	204484	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	218557	47.958	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.920%
35) Dibromofluoromethane	5.385	113	161795	48.447	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	8.647	98	603530	48.961	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.920%
62) 4-Bromofluorobenzene	11.079	95	232246	48.687	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.380%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040323\
Data File : VX034919.D
Acq On : 03 Apr 2023 10:06
Operator : JC/MD
Sample : VX0403WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0403WBL01


Quant Time: Apr 03 21:32:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX034919.D
Acq: 03 Apr 2023 10:06

Instrument :
MSVOA_X
ClientSampleId :
VX0403WBL01







Tgt Ion:168 Resp: 297985
Ion Ratio Lower Upper
168 100
99 64.0 45.7 68.5



#33
1,2-Dichloroethane-d4
Concen: 47.958 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.006 min
Lab File: VX034919.D
Acq: 03 Apr 2023 10:06

Tgt Ion: 65 Resp: 218557
Ion Ratio Lower Upper
65 100
67 51.0 0.0 104.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX034919.D
Acq: 03 Apr 2023 10:06




Instrument :
MSVOA_X
ClientSampleId :
VX0403WBL01


Tgt Ion:114 Resp: 504919
Ion Ratio Lower Upper
114 100
63 22.3 0.0 42.2
88 16.7 0.0 33.2



#35
Dibromofluoromethane
Concen: 48.447 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX034919.D
Acq: 03 Apr 2023 10:06

Tgt Ion:113 Resp: 161795
Ion Ratio Lower Upper
113 100
111 101.7 83.7 125.5
192 18.2 15.4 23.2









#50
Toluene-d8
Concen: 48.961 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX034919.D
Acq: 03 Apr 2023 10:06




Instrument :
MSVOA_X
ClientSampleId :
VX0403WBL01


Tgt Ion: 98 Resp: 603530
Ion Ratio Lower Upper
98 100
100 66.5 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 48.687 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX034919.D
Acq: 03 Apr 2023 10:06

Tgt Ion: 95 Resp: 232246
Ion Ratio Lower Upper
95 100
174 70.7 0.0 147.4
176 67.6 0.0 140.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX034919.D
Acq: 03 Apr 2023 10:06

Instrument :
MSVOA_X
ClientSampleId :
VX0403WBL01

Tgt Ion:117 Resp: 449098
Ion Ratio Lower Upper
117 100
82 57.9 45.0 67.4
119 33.0 26.0 39.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX034919.D
Acq: 03 Apr 2023 10:06

Tgt Ion:152 Resp: 204484
Ion Ratio Lower Upper
152 100
115 63.3 45.0 135.0
150 155.6 0.0 352.4

