

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040324\
 Data File : VX040862.D
 Acq On : 03 Apr 2024 10:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Apr 03 16:37:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	52239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	87562	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	82342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	43097	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	43574	46.488	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.980%
35) Dibromofluoromethane	5.379	113	29734	47.943	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.880%
50) Toluene-d8	8.647	98	105966	47.283	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.560%
62) 4-Bromofluorobenzene	11.079	95	42274	47.235	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.480%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	32132	45.380	ug/l	100
3) Chloromethane	1.301	50	31536	51.555	ug/l	99
4) Vinyl Chloride	1.374	62	33281	52.708	ug/l	100
5) Bromomethane	1.599	94	22114	56.928	ug/l	98
6) Chloroethane	1.685	64	20980	51.545	ug/l	96
7) Trichlorofluoromethane	1.886	101	57542	47.607	ug/l	99
8) Diethyl Ether	2.136	74	18432	47.190	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	29022	46.575	ug/l	98
10) Methyl Iodide	2.453	142	35089	46.421	ug/l	99
11) Tert butyl alcohol	2.953	59	37418	218.658	ug/l #	92
12) 1,1-Dichloroethene	2.319	96	26393	45.161	ug/l	96
13) Acrolein	2.233	56	20548	219.972	ug/l	99
14) Allyl chloride	2.660	41	49844	47.610	ug/l	90
15) Acrylonitrile	3.062	53	85048	235.595	ug/l	97
16) Acetone	2.374	43	85269	234.288	ug/l	96
17) Carbon Disulfide	2.514	76	75093	44.416	ug/l	98
18) Methyl Acetate	2.703	43	38415	44.490	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	96802	47.057	ug/l	99
20) Methylene Chloride	2.788	84	29560	43.827	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	27591	44.698	ug/l	98
22) Diisopropyl ether	3.757	45	98867	48.526	ug/l	97
23) Vinyl Acetate	3.715	43	482134	246.898	ug/l	99
24) 1,1-Dichloroethane	3.605	63	54895	47.160	ug/l	98
25) 2-Butanone	4.550	43	126746	234.435	ug/l	99
26) 2,2-Dichloropropane	4.471	77	47469	47.500	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	33976	47.665	ug/l	99
28) Bromochloromethane	4.897	49	24644	47.086	ug/l	98
29) Tetrahydrofuran	4.995	42	81892	233.043	ug/l	99
30) Chloroform	5.086	83	60191	47.508	ug/l	98
31) Cyclohexane	5.471	56	45233	45.477	ug/l	99
32) 1,1,1-Trichloroethane	5.379	97	52998	47.073	ug/l	97
36) 1,1-Dichloropropene	5.690	75	41177	46.726	ug/l	98
37) Ethyl Acetate	4.708	43	50562	46.725	ug/l	97
38) Carbon Tetrachloride	5.678	117	45742	46.888	ug/l	96
39) Methylcyclohexane	7.379	83	46726	44.377	ug/l	99
40) Benzene	6.037	78	115471	47.211	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040324\
 Data File : VX040862.D
 Acq On : 03 Apr 2024 10:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Apr 03 16:37:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	27244	49.024	ug/l	96
42) 1,2-Dichloroethane	6.080	62	50874	47.890	ug/l	96
43) Isopropyl Acetate	6.330	43	81813	48.956	ug/l	96
44) Trichloroethene	7.123	130	29041	47.541	ug/l	97
45) 1,2-Dichloropropane	7.428	63	29651	50.783	ug/l	100
46) Dibromomethane	7.574	93	23486	49.076	ug/l	98
47) Bromodichloromethane	7.818	83	48353	48.561	ug/l	95
48) Methyl methacrylate	7.690	41	41080	50.199	ug/l	95
49) 1,4-Dioxane	7.653	88	15296	979.493	ug/l #	96
51) 4-Methyl-2-Pentanone	8.568	43	267929	241.707	ug/l	95
52) Toluene	8.714	92	72567	47.020	ug/l	94
53) t-1,3-Dichloropropene	8.976	75	47920	46.790	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	48544	49.168	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	29224	48.522	ug/l	95
56) Ethyl methacrylate	9.110	69	50800	49.120	ug/l #	92
57) 1,3-Dichloropropane	9.305	76	51164	48.573	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	124638	248.597	ug/l	97
59) 2-Hexanone	9.427	43	212651	239.200	ug/l	95
60) Dibromochloromethane	9.519	129	35694	49.644	ug/l	100
61) 1,2-Dibromoethane	9.604	107	31925	49.131	ug/l	98
64) Tetrachloroethene	9.269	164	24976	45.836	ug/l	97
65) Chlorobenzene	10.079	112	82587	45.861	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	28188	47.157	ug/l	99
67) Ethyl Benzene	10.189	91	148619	46.277	ug/l	96
68) m/p-Xylenes	10.299	106	108467	92.826	ug/l	90
69) o-Xylene	10.640	106	52018	47.254	ug/l	90
70) Styrene	10.653	104	91549	47.992	ug/l	97
71) Bromoform	10.799	173	24717	46.860	ug/l #	98
73) Isopropylbenzene	10.963	105	143204	47.576	ug/l	100
74) N-amyl acetate	10.842	43	75266	49.368	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.207	83	49250	47.852	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	57582	62.657	ug/l	89
77) Bromobenzene	11.195	156	33936	48.062	ug/l	98
78) n-propylbenzene	11.305	91	176933	47.034	ug/l	98
79) 2-Chlorotoluene	11.360	91	103065	45.909	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	124583	48.461	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.018	75	16373	48.068	ug/l	97
82) 4-Chlorotoluene	11.451	91	127905	47.720	ug/l	98
83) tert-Butylbenzene	11.713	119	120809	47.369	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	126256	48.752	ug/l	99
85) sec-Butylbenzene	11.890	105	155421	47.983	ug/l	99
86) p-Isopropyltoluene	12.006	119	129063	47.507	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	66056	46.282	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	67978	44.887	ug/l	99
89) n-Butylbenzene	12.329	91	129247	48.156	ug/l	97
90) Hexachloroethane	12.536	117	23129	48.872	ug/l	93
91) 1,2-Dichlorobenzene	12.335	146	63988	45.747	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	13270	50.177	ug/l	95
93) 1,2,4-Trichlorobenzene	13.585	180	44194	45.576	ug/l	96
94) Hexachlorobutadiene	13.725	225	16320	43.935	ug/l	98
95) Naphthalene	13.774	128	151616	48.560	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	43842	47.393	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040324\
Data File : VX040862.D
Acq On : 03 Apr 2024 10:13
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Apr 03 16:37:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 52239
Ion Ratio Lower Upper
168 100
99 67.1 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 45.380 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion: 85 Resp: 32132
Ion Ratio Lower Upper
85 100
87 33.3 16.7 50.1








#3
Chloromethane
Concen: 51.555 ug/l
RT: 1.301 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 50 Resp: 31536
Ion Ratio Lower Upper
50 100
52 32.5 25.5 38.3




#4
Vinyl Chloride
Concen: 52.708 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 62 Resp: 33281
Ion Ratio Lower Upper
62 100
64 31.9 25.4 38.0



Abundance Scan 84 (1.599 min): VX040845.D\data.ms (-80)



#5

Bromomethane

Concen: 56.928 ug/l

RT: 1.599 min Scan# 84

Delta R.T. 0.000 min

Lab File: VX040862.D

Acq: 03 Apr 2024 10:13

Instrument :

MSVOA_X

ClientSampleId :


Tgt Ion: 94 Resp: 22114

Ion Ratio Lower Upper


94 100

96 91.4 74.6 111.8


Abundance Scan 84 (1.599 min): VX040862.D\data.ms




Abundance



Abundance Scan 84 (1.599 min): VX040862.D\data.ms (-35)



Abundance Scan 98 (1.684 min): VX040845.D\data.ms (-93)



#6

Chloroethane

Concen: 51.545 ug/l

RT: 1.685 min Scan# 98

Delta R.T. 0.000 min

Lab File: VX040862.D

Acq: 03 Apr 2024 10:13


Tgt Ion: 64 Resp: 20980

Ion Ratio Lower Upper


64 100

66 29.3 25.1 37.7


Abundance Scan 98 (1.685 min): VX040862.D\data.ms




Abundance



Abundance Scan 98 (1.685 min): VX040862.D\data.ms (-49)









#7
Trichlorofluoromethane
Concen: 47.607 ug/l
RT: 1.886 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 57542
Ion Ratio Lower Upper
101 100
103 65.4 51.4 77.2



#8
Diethyl Ether
Concen: 47.190 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 74 Resp: 18432
Ion Ratio Lower Upper
74 100
45 106.8 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 46.575 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 29022
Ion Ratio Lower Upper
101 100
85 47.4 37.0 55.6
151 72.0 58.6 87.8



#10
Methyl Iodide
Concen: 46.421 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion:142 Resp: 35089
Ion Ratio Lower Upper
142 100
127 51.3 41.6 62.4
141 13.8 11.6 17.4








#11
Tert butyl alcohol
Concen: 218.658 ug/l
RT: 2.953 min Scan# 307
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 59 Resp: 37418
Ion Ratio Lower Upper
59 100
57 10.3 6.0 9.0#



#12
1,1-Dichloroethene
Concen: 45.161 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 96 Resp: 26393
Ion Ratio Lower Upper
96 100
61 188.3 144.6 216.8
98 63.2 50.4 75.6









#13
Acrolein
Concen: 219.972 ug/l
RT: 2.233 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 20548
Ion Ratio Lower Upper
56 100
55 72.5 57.1 85.7



#14
Allyl chloride
Concen: 47.610 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 41 Resp: 49844
Ion Ratio Lower Upper
41 100
39 80.3 55.8 83.6
76 33.0 24.3 36.5









#15
Acrylonitrile
Concen: 235.595 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 85048
Ion Ratio Lower Upper
53 100
52 85.0 66.0 99.0
51 38.8 29.8 44.6



#16
Acetone
Concen: 234.288 ug/l
RT: 2.374 min Scan# 211
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 43 Resp: 85269
Ion Ratio Lower Upper
43 100
58 25.5 21.9 32.9









#17
Carbon Disulfide
Concen: 44.416 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 75093
Ion Ratio Lower Upper
76 100
78 9.4 7.0 10.6



#18
Methyl Acetate
Concen: 44.490 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 43 Resp: 38415
Ion Ratio Lower Upper
43 100
74 22.4 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 47.057 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 96802
Ion Ratio Lower Upper
73 100
57 21.2 17.3 25.9



#20
Methylene Chloride
Concen: 43.827 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion: 84 Resp: 29560
Ion Ratio Lower Upper
84 100
49 136.2 103.6 155.4
51 42.6 33.8 50.6
86 63.7 49.4 74.0








#21
trans-1,2-Dichloroethene
Concen: 44.698 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 27591
Ion Ratio Lower Upper
96 100
61 150.1 119.2 178.8
98 63.4 48.7 73.1




#22
Diisopropyl ether
Concen: 48.526 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13



Tgt Ion: 45 Resp: 98867
Ion Ratio Lower Upper
45 100
43 83.7 69.7 104.5
87 26.6 20.8 31.2
59 11.1 9.0 13.4









#23
Vinyl Acetate
Concen: 246.898 ug/l
RT: 3.715 min Scan# 413
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 482134
Ion Ratio Lower Upper
43 100
86 9.3 7.8 11.8



#24
1,1-Dichloroethane
Concen: 47.160 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 63 Resp: 54895
Ion Ratio Lower Upper
63 100
98 5.6 3.3 9.8
100 4.3 2.1 6.5









#25
2-Butanone
Concen: 234.435 ug/l
RT: 4.550 min Scan# 51
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 126746
Ion Ratio Lower Upper
43 100
72 24.9 19.4 29.2



#26
2,2-Dichloropropane
Concen: 47.500 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion: 77 Resp: 47469
Ion Ratio Lower Upper
77 100
97 20.9 10.8 32.3








#27
cis-1,2-Dichloroethene
Concen: 47.665 ug/l
RT: 4.489 min Scan# 511
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 33976
Ion Ratio Lower Upper
96 100
61 156.2 0.0 312.2
98 62.7 0.0 129.0



#28
Bromochloromethane
Concen: 47.086 ug/l
RT: 4.897 min Scan# 625
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 49 Resp: 24644
Ion Ratio Lower Upper
49 100
129 1.8 0.0 3.8
130 74.0 57.5 86.3









#29
Tetrahydrofuran
Concen: 233.043 ug/l
RT: 4.995 min Scan# 641
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 42 Resp: 81892




Ion	Ratio	Lower	Upper
42	100		
72	42.2	34.2	51.4
71	40.1	32.2	48.2




#30
Chloroform
Concen: 47.508 ug/l
RT: 5.086 min Scan# 656
Delta R.T. -0.012 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 83 Resp: 60191


Ion	Ratio	Lower	Upper
83	100		
85	64.3	50.3	75.5








#31
Cyclohexane
Concen: 45.477 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 45233
Ion Ratio Lower Upper
56 100
69 28.6 23.4 35.2
84 81.3 65.9 98.9



#32
1,1,1-Trichloroethane
Concen: 47.073 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 97 Resp: 52998
Ion Ratio Lower Upper
97 100
99 64.9 51.1 76.7
61 48.8 36.4 54.6









#33
1,2-Dichloroethane-d4
Concen: 46.488 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 65 Resp: 43574
Ion Ratio Lower Upper
65 100
67 48.6 0.0 100.2



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 114 Resp: 87562
Ion Ratio Lower Upper
114 100
63 22.0 0.0 45.0
88 18.2 0.0 36.6





#35

Dibromofluoromethane

Concen: 47.943 ug/l

RT: 5.379 min Scan# 705

Delta R.T. -0.006 min

Lab File: VX040862.D





Acq: 03 Apr 2024 10:13

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:113	Resp:	29734
Ion	Ratio	Lower Upper
113	100	
111	104.3	82.9 124.3
192	17.9	13.3 19.9



#36

1,1-Dichloropropene

Concen: 46.726 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.000 min

Lab File: VX040862.D

Acq: 03 Apr 2024 10:13


Tgt Ion: 75	Resp:	41177
Ion	Ratio	Lower Upper
75	100	
110	32.6	17.0 50.9
77	30.5	24.7 37.1








#37
Ethyl Acetate
Concen: 46.725 ug/l
RT: 4.708 min Scan# 594
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 50562
Ion Ratio Lower Upper
43 100
61 12.5 10.6 16.0
70 9.2 8.3 12.5



#38
Carbon Tetrachloride
Concen: 46.888 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion: 117 Resp: 45742
Ion Ratio Lower Upper
117 100
119 93.2 78.2 117.2
121 29.7 24.4 36.6








#39
Methylcyclohexane
Concen: 44.377 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83 Resp: 46726
Ion Ratio Lower Upper
83 100
55 81.1 64.6 97.0
98 45.4 36.7 55.1



#40
Benzene
Concen: 47.211 ug/l
RT: 6.037 min Scan# 812
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion: 78 Resp: 115471
Ion Ratio Lower Upper
78 100
77 23.1 19.3 28.9








#41
Methacrylonitrile
Concen: 49.024 ug/l
RT: 4.910 min Scan# 61
Delta R.T. -0.012 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41 Resp: 27244
Ion Ratio Lower Upper
41 100
39 60.4 45.2 67.8
67 66.4 52.6 78.8
52 33.0 22.7 34.1



#42
1,2-Dichloroethane
Concen: 47.890 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion: 62 Resp: 50874
Ion Ratio Lower Upper
62 100
98 6.7 0.0 16.0








#43
Isopropyl Acetate
Concen: 48.956 ug/l
RT: 6.330 min Scan# 860
Delta R.T. -0.012 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 81813
Ion Ratio Lower Upper
43 100
61 16.4 15.2 22.8
87 11.6 9.8 14.6



#44
Trichloroethene
Concen: 47.541 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion:130 Resp: 29041
Ion Ratio Lower Upper
130 100
95 99.1 0.0 204.4









#45
1,2-Dichloropropane
Concen: 50.783 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 29651
Ion Ratio Lower Upper
63 100
65 30.6 24.3 36.5



#46
Dibromomethane
Concen: 49.076 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 93 Resp: 23486
Ion Ratio Lower Upper
93 100
95 81.1 66.6 100.0
174 88.3 69.4 104.0





#47

Bromodichloromethane

Concen: 48.561 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.006 min


Lab File: VX040862.D

Acq: 03 Apr 2024 10:13

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 83 Resp: 48353

Ion Ratio Lower Upper

83 100

85 66.4 49.8 74.6

127 8.2 7.1 10.7



#48

Methyl methacrylate

Concen: 50.199 ug/l

RT: 7.690 min Scan# 1083

Delta R.T. 0.000 min

Lab File: VX040862.D

Acq: 03 Apr 2024 10:13




Tgt Ion: 41 Resp: 41080


Ion Ratio Lower Upper

41 100

69 73.4 59.2 88.8

39 67.8 47.7 71.5





#49

1,4-Dioxane

Concen: 979.493 ug/l

RT: 7.653 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX040862.D

Acq: 03 Apr 2024 10:13

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 15296

Ion Ratio Lower Upper

88 100

43 40.6 0.0 0.0#

58 71.6 60.1 90.1



#50

Toluene-d8

Concen: 47.283 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX040862.D




Acq: 03 Apr 2024 10:13


Tgt Ion: 98 Resp: 105966

Ion Ratio Lower Upper

98 100

100 62.5 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 241.707 ug/l
RT: 8.568 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 267929
Ion Ratio Lower Upper
43 100
58 35.3 30.4 45.6



#52
Toluene
Concen: 47.020 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 92 Resp: 72567
Ion Ratio Lower Upper
92 100
91 176.4 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 46.790 ug/l

RT: 8.976 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX040862.D

Acq: 03 Apr 2024 10:13

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 47920

Ion Ratio Lower Upper

75 100

77 32.0 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 49.168 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. -0.006 min

Lab File: VX040862.D

Acq: 03 Apr 2024 10:13




Tgt Ion: 75 Resp: 48544


Ion Ratio Lower Upper

75 100

77 31.9 25.3 37.9

39 61.4 46.7 70.1









#55
1,1,2-Trichloroethane
Concen: 48.522 ug/l
RT: 9.147 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 29224




Ion	Ratio	Lower	Upper
97	100		
83	87.4	65.8	98.8
85	55.9	41.6	62.4
99	61.6	47.8	71.8




#56
Ethyl methacrylate
Concen: 49.120 ug/l
RT: 9.110 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 69 Resp: 50800

Ion	Ratio	Lower	Upper
69	100		
41	75.4	56.8	85.2
39	48.2	32.0	48.0









#57
1,3-Dichloropropane
Concen: 48.573 ug/l
RT: 9.305 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 51164
Ion Ratio Lower Upper
76 100
78 31.6 26.1 39.1



#58
2-Chloroethyl Vinyl ether
Concen: 248.597 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 63 Resp: 124638
Ion Ratio Lower Upper
63 100
106 26.5 22.5 33.7









#59
2-Hexanone
Concen: 239.200 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 212651
Ion Ratio Lower Upper
43 100
58 49.1 26.2 78.5



#60
Dibromochloromethane
Concen: 49.644 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion:129 Resp: 35694
Ion Ratio Lower Upper
129 100
127 77.8 38.8 116.4









#61
1,2-Dibromoethane
Concen: 49.131 ug/l
RT: 9.604 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 31925
Ion Ratio Lower Upper
107 100
109 91.7 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 47.235 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion: 95 Resp: 42274
Ion Ratio Lower Upper
95 100
174 73.1 0.0 135.6
176 69.9 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 82342
Ion Ratio Lower Upper
117 100
82 59.6 46.2 69.4
119 32.1 25.5 38.3



#64
Tetrachloroethene
Concen: 45.836 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion:164 Resp: 24976
Ion Ratio Lower Upper
164 100
166 123.8 102.7 154.1
129 102.6 80.4 120.6
131 96.5 76.5 114.7








#65
Chlorobenzene
Concen: 45.861 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:112 Resp: 82587
Ion Ratio Lower Upper
112 100
114 33.0 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 47.157 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion:131 Resp: 28188
Ion Ratio Lower Upper
131 100
133 96.3 47.1 141.4
119 64.5 32.2 96.6





#67

Ethyl Benzene

Concen: 46.277 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.006 min

Lab File: VX040862.D

Acq: 03 Apr 2024 10:13

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 148619

Ion Ratio Lower Upper

91 100

106 28.9 24.9 37.3



#68

m/p-Xylenes

Concen: 92.826 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX040862.D




Acq: 03 Apr 2024 10:13


Tgt Ion:106 Resp: 108467

Ion Ratio Lower Upper

106 100

91 222.9 165.5 248.3









#69
o-Xylene
Concen: 47.254 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 52018
Ion Ratio Lower Upper
106 100
91 233.3 108.9 326.8



#70
Styrene
Concen: 47.992 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion:104 Resp: 91549
Ion Ratio Lower Upper
104 100
78 57.8 43.6 65.4
103 55.5 43.6 65.4








#71
Bromoform
Concen: 46.860 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:173 Resp: 24717
Ion Ratio Lower Upper
173 100
175 47.2 24.1 72.4
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion:152 Resp: 43097
Ion Ratio Lower Upper
152 100
115 91.3 43.3 129.9
150 173.2 0.0 345.0









#73
Isopropylbenzene
Concen: 47.576 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 143204
Ion Ratio Lower Upper
105 100
120 25.5 12.9 38.6



#74
N-ethyl acetate
Concen: 49.368 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 43 Resp: 75266
Ion Ratio Lower Upper
43 100
70 36.8 32.9 49.3
55 23.0 19.3 28.9
61 21.8 18.8 28.2









#75
1,1,2,2-Tetrachloroethane
Concen: 47.852 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 49250
Ion Ratio Lower Upper
83 100
131 9.9 5.3 15.8
85 63.3 32.0 96.2



#76
1,2,3-Trichloropropane
Concen: 62.657 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 75 Resp: 57582
Ion Ratio Lower Upper
75 100
77 31.4 19.1 57.5









#77
Bromobenzene
Concen: 48.062 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 33936
Ion Ratio Lower Upper
156 100
77 179.6 88.4 265.3
158 97.0 48.0 144.0



#78
n-propylbenzene
Concen: 47.034 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion: 91 Resp: 176933
Ion Ratio Lower Upper
91 100
120 21.8 11.5 34.4









#79
2-Chlorotoluene
Concen: 45.909 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 103065
Ion Ratio Lower Upper
91 100
126 31.9 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 48.461 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion:105 Resp: 124583
Ion Ratio Lower Upper
105 100
120 45.5 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 48.068 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 16373
Ion Ratio Lower Upper
75 100
53 117.0 90.5 135.7
89 47.4 36.9 55.3



#82
4-Chlorotoluene
Concen: 47.720 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion: 91 Resp: 127905
Ion Ratio Lower Upper
91 100
126 27.3 14.2 42.6








#83
tert-Butylbenzene
Concen: 47.369 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:119 Resp: 120809
Ion Ratio Lower Upper
119 100
91 66.3 31.2 93.6
134 22.3 11.7 35.0



#84
1,2,4-Trimethylbenzene
Concen: 48.752 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion:105 Resp: 126256
Ion Ratio Lower Upper
105 100
120 41.8 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 47.983 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX040862.D
 Acq: 03 Apr 2024 10:13




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 155421
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 47.507 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX040862.D
 Acq: 03 Apr 2024 10:13


Tgt Ion:119 Resp: 129063
 Ion Ratio Lower Upper
 119 100
 134 26.4 12.8 38.3
 91 28.6 13.1 39.3








#87
1,3-Dichlorobenzene
Concen: 46.282 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:146 Resp: 66056
Ion Ratio Lower Upper
146 100
111 43.1 22.1 66.1
148 64.4 31.8 95.3




#88
1,4-Dichlorobenzene
Concen: 44.887 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13



Tgt Ion:146 Resp: 67978
Ion Ratio Lower Upper
146 100
111 42.4 21.3 63.7
148 64.4 31.8 95.4





#89

n-Butylbenzene

Concen: 48.156 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX040862.D

Acq: 03 Apr 2024 10:13





Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 91 Resp: 129247

Ion	Ratio	Lower	Upper
91	100		
92	51.9	27.4	82.0
134	24.0	12.6	37.8



#90

Hexachloroethane

Concen: 48.872 ug/l

RT: 12.536 min Scan# 1878




Delta R.T. 0.000 min


Lab File: VX040862.D

Acq: 03 Apr 2024 10:13

Tgt Ion: 117 Resp: 23129


Ion	Ratio	Lower	Upper
117	100		
201	71.5	32.9	98.6








#91
1,2-Dichlorobenzene
Concen: 45.747 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 63988
Ion Ratio Lower Upper
146 100
111 44.9 22.4 67.0
148 64.3 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 50.177 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.006 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Tgt Ion: 75 Resp: 13270
Ion Ratio Lower Upper
75 100
155 67.1 36.1 108.5
157 86.4 44.9 134.7








#93
1,2,4-Trichlorobenzene
Concen: 45.576 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:180 Resp: 44194
Ion Ratio Lower Upper
180 100
182 93.7 48.3 144.8
145 32.8 18.0 54.0




#94
Hexachlorobutadiene
Concen: 43.935 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13




Tgt Ion:225 Resp: 16320
Ion Ratio Lower Upper
225 100
223 63.0 31.1 93.2
227 67.1 32.5 97.5








#95
Naphthalene
Concen: 48.560 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 151616
Ion Ratio Lower Upper
128 100
127 13.1 10.3 15.5
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 47.393 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX040862.D
Acq: 03 Apr 2024 10:13

Tgt Ion:180 Resp: 43842
Ion Ratio Lower Upper
180 100
182 93.3 46.9 140.6
145 34.0 17.8 53.5

