

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040324\
 Data File : VX040868.D
 Acq On : 03 Apr 2024 12:46
 Operator : JC/MD
 Sample : P1964-01DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RSB-02-17-040224DL

Quant Time: Apr 03 16:39:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 04/03/2024
 Supervised By :Semsettin Yesilyurt 04/03/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	43982	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	78570	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	79564	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	42049	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	38012	48.168	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.340%		
35) Dibromofluoromethane	5.385	113	26710	47.995	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.000%		
50) Toluene-d8	8.647	98	100386	49.920	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.840%		
62) 4-Bromofluorobenzene	11.079	95	41981	52.276	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.560%		
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	19858	37.354	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	25754	42.913	ug/l	90
30) Chloroform	5.086	83	1663m	1.559	ug/l	
40) Benzene	6.050	78	1058	0.482	ug/l #	88
65) Chlorobenzene	10.079	112	1680m	0.965	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040324\
Data File : VX040868.D
Acq On : 03 Apr 2024 12:46
Operator : JC/MD
Sample : P1964-01DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RSB-02-17-040224DL

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/03/2024
Supervised By :Semsettin Yesilyurt 04/03/2024

Quant Time: Apr 03 16:39:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

