

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040324\
 Data File : VX040871.D
 Acq On : 03 Apr 2024 14:01
 Operator : JC/MD
 Sample : P1964-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-040224

Quant Time: Apr 03 16:39:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	43579	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	77910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	78027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	40025	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38166	48.810	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.620%		
35) Dibromofluoromethane	5.385	113	27405	49.661	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.320%		
50) Toluene-d8	8.647	98	99094	49.695	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.380%		
62) 4-Bromofluorobenzene	11.079	95	40303	50.612	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.220%		
Target Compounds						
16) Acetone	2.380	43	1994	6.568	ug/l	99
20) Methylene Chloride	2.794	84	1102	1.959	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040324\
Data File : VX040871.D
Acq On : 03 Apr 2024 14:01
Operator : JC/MD
Sample : P1964-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-040224

Quant Time: Apr 03 16:39:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

