

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040324\  
 Data File : VX040872.D  
 Acq On : 03 Apr 2024 15:56  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 LabSampleId :  
 VSTDCCC050

Quant Time: Apr 03 16:39:40 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040124W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 02 05:40:55 2024  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|--------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0    | 92    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.678 | 0.669 | 1.3    | 90    | 0.00     |
| 3 P   | Chloromethane               | 0.585 | 0.643 | -9.9   | 100   | 0.00     |
| 4 C   | Vinyl Chloride              | 0.604 | 0.690 | -14.2# | 106   | 0.00     |
| 5 T   | Bromomethane                | 0.372 | 0.470 | -26.3# | 104   | 0.00     |
| 6 T   | Chloroethane                | 0.390 | 0.433 | -11.0  | 105   | 0.00     |
| 7 T   | Trichlorofluoromethane      | 1.157 | 1.213 | -4.8   | 97    | 0.00     |
| 8 T   | Diethyl Ether               | 0.374 | 0.383 | -2.4   | 98    | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.596 | 0.603 | -1.2   | 95    | 0.00     |
| 10 T  | Methyl Iodide               | 0.723 | 0.722 | 0.1    | 90    | 0.00     |
| 11 T  | Tert butyl alcohol          | 0.164 | 0.150 | 8.5    | 84    | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 0.559 | 0.547 | 2.1#   | 93    | 0.00     |
| 13 T  | Acrolein                    | 0.089 | 0.090 | -1.1   | 92    | 0.00     |
| 14 T  | Allyl chloride              | 1.002 | 1.036 | -3.4   | 94    | 0.00     |
| 15 T  | Acrylonitrile               | 0.346 | 0.342 | 1.2    | 90    | 0.00     |
| 16 T  | Acetone                     | 0.348 | 0.355 | -2.0   | 97    | 0.00     |
| 17 T  | Carbon Disulfide            | 1.618 | 1.547 | 4.4    | 94    | 0.00     |
| 18 T  | Methyl Acetate              | 0.826 | 0.772 | 6.5    | 85    | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 1.969 | 1.992 | -1.2   | 94    | 0.00     |
| 20 T  | Methylene Chloride          | 0.646 | 0.609 | 5.7    | 95    | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 0.591 | 0.581 | 1.7    | 97    | 0.00     |
| 22 T  | Diisopropyl ether           | 1.950 | 2.008 | -3.0   | 96    | 0.00     |
| 23 T  | Vinyl Acetate               | 1.869 | 1.986 | -6.3   | 96    | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 1.114 | 1.117 | -0.3   | 94    | 0.00     |
| 25 T  | 2-Butanone                  | 0.517 | 0.516 | 0.2    | 91    | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 0.957 | 0.999 | -4.4   | 97    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 0.682 | 0.685 | -0.4   | 95    | 0.00     |
| 28 T  | Bromochloromethane          | 0.501 | 0.497 | 0.8    | 96    | -0.01    |
| 29 T  | Tetrahydrofuran             | 0.336 | 0.327 | 2.7    | 89    | 0.00     |
| 30 C  | Chloroform                  | 1.213 | 1.263 | -4.1#  | 97    | -0.01    |
| 31 T  | Cyclohexane                 | 0.952 | 0.940 | 1.3    | 91    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 1.078 | 1.110 | -3.0   | 93    | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 0.897 | 0.910 | -1.4   | 99    | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0    | 92    | 0.00     |
| 35 S  | Dibromofluoromethane        | 0.354 | 0.382 | -7.9   | 100   | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 0.503 | 0.504 | -0.2   | 94    | 0.00     |
| 37 T  | Ethyl Acetate               | 0.618 | 0.609 | 1.5    | 90    | 0.00     |
| 38 T  | Carbon Tetrachloride        | 0.557 | 0.592 | -6.3   | 96    | 0.00     |
| 39 T  | Methylcyclohexane           | 0.601 | 0.595 | 1.0    | 92    | 0.00     |
| 40 TM | Benzene                     | 1.397 | 1.437 | -2.9   | 95    | 0.00     |
| 41 T  | Methacrylonitrile           | 0.317 | 0.347 | -9.5   | 96    | -0.01    |
| 42 TM | 1,2-Dichloroethane          | 0.607 | 0.638 | -5.1   | 96    | 0.00     |
| 43 T  | Isopropyl Acetate           | 0.954 | 0.993 | -4.1   | 93    | -0.01    |
| 44 TM | Trichloroethene             | 0.349 | 0.359 | -2.9   | 95    | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 0.333 | 0.358 | -7.5#  | 96    | 0.00     |
| 46 T  | Dibromomethane              | 0.273 | 0.293 | -7.3   | 97    | 0.00     |
| 47 T  | Bromodichloromethane        | 0.569 | 0.600 | -5.4   | 96    | 0.00     |
| 48 T  | Methyl methacrylate         | 0.467 | 0.506 | -8.4   | 93    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040324\  
 Data File : VX040872.D  
 Acq On : 03 Apr 2024 15:56  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 LabSampleId :  
 VSTDCCC050

Quant Time: Apr 03 16:39:40 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040124W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 02 05:40:55 2024  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 49 T  | 1,4-Dioxane                 | 0.009 | 0.009 | 0.0   | 83    | 0.00     |
| 50 S  | Toluene-d8                  | 1.280 | 1.356 | -5.9  | 99    | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 0.633 | 0.658 | -3.9  | 91    | 0.00     |
| 52 CM | Toluene                     | 0.881 | 0.930 | -5.6# | 96    | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 0.540 | 0.598 | -10.7 | 94    | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 0.564 | 0.620 | -9.9  | 97    | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 0.344 | 0.361 | -4.9  | 93    | 0.00     |
| 56 T  | Ethyl methacrylate          | 0.591 | 0.628 | -6.3  | 92    | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 0.601 | 0.631 | -5.0  | 97    | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.286 | 0.305 | -6.6  | 96    | 0.00     |
| 59 T  | 2-Hexanone                  | 0.508 | 0.528 | -3.9  | 92    | 0.00     |
| 60 T  | Dibromochloromethane        | 0.411 | 0.441 | -7.3  | 96    | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 0.371 | 0.387 | -4.3  | 95    | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 0.511 | 0.550 | -7.6  | 102   | 0.00     |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0   | 93    | 0.00     |
| 64 T  | Tetrachloroethene           | 0.331 | 0.338 | -2.1  | 97    | 0.00     |
| 65 PM | Chlorobenzene               | 1.094 | 1.096 | -0.2  | 97    | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.363 | 0.378 | -4.1  | 95    | 0.00     |
| 67 C  | Ethyl Benzene               | 1.950 | 2.012 | -3.2# | 96    | 0.00     |
| 68 T  | m/p-Xylenes                 | 0.710 | 0.729 | -2.7  | 96    | 0.00     |
| 69 T  | o-Xylene                    | 0.668 | 0.696 | -4.2  | 96    | 0.00     |
| 70 T  | Styrene                     | 1.158 | 1.242 | -7.3  | 96    | 0.00     |
| 71 P  | Bromoform                   | 0.302 | 0.324 | -7.3  | 94    | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0   | 95    | 0.00     |
| 73 T  | Isopropylbenzene            | 3.492 | 3.541 | -1.4  | 95    | 0.00     |
| 74 T  | N-amyl acetate              | 1.769 | 1.838 | -3.9  | 95    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 1.194 | 1.188 | 0.5   | 94    | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 1.066 | 1.063 | 0.3   | 98    | 0.00     |
| 77 T  | Bromobenzene                | 0.819 | 0.833 | -1.7  | 94    | 0.00     |
| 78 T  | n-propylbenzene             | 4.364 | 4.434 | -1.6  | 97    | 0.00     |
| 79 T  | 2-Chlorotoluene             | 2.605 | 2.612 | -0.3  | 99    | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 2.983 | 3.078 | -3.2  | 96    | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.395 | 0.389 | 1.5   | 89    | 0.00     |
| 82 T  | 4-Chlorotoluene             | 3.110 | 3.173 | -2.0  | 98    | 0.00     |
| 83 T  | tert-Butylbenzene           | 2.959 | 2.981 | -0.7  | 96    | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 3.005 | 3.083 | -2.6  | 95    | 0.00     |
| 85 T  | sec-Butylbenzene            | 3.758 | 3.862 | -2.8  | 97    | 0.00     |
| 86 T  | p-Isopropyltoluene          | 3.152 | 3.262 | -3.5  | 96    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 1.656 | 1.639 | 1.0   | 97    | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 1.757 | 1.698 | 3.4   | 98    | 0.00     |
| 89 T  | n-Butylbenzene              | 3.114 | 3.255 | -4.5  | 97    | 0.00     |
| 90 T  | Hexachloroethane            | 0.549 | 0.583 | -6.2  | 99    | 0.00     |
| 91 T  | 1,2-Dichlorobenzene         | 1.623 | 1.572 | 3.1   | 95    | 0.00     |
| 92 T  | 1,2-Dibromo-3-Chloropropane | 0.307 | 0.309 | -0.7  | 88    | 0.00     |
| 93 T  | 1,2,4-Trichlorobenzene      | 1.125 | 1.091 | 3.0   | 96    | 0.00     |
| 94 T  | Hexachlorobutadiene         | 0.431 | 0.421 | 2.3   | 97    | 0.00     |
| 95 T  | Naphthalene                 | 3.622 | 3.681 | -1.6  | 94    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040324\  
Data File : VX040872.D  
Acq On : 03 Apr 2024 15:56  
Operator : JC/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
LabSampleId :  
VSTDCCC050

Quant Time: Apr 03 16:39:40 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040124W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 02 05:40:55 2024  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 96 T | 1,2,3-Trichlorobenzene | 1.073 | 1.065 | 0.7  | 96    | 0.00     |

(#) = Out of Range                      SPCC's out = 0    CCC's out = 6