

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
 Data File : VX045562.D
 Acq On : 03 Apr 2025 11:45
 Operator : JC/MD
 Sample : Q1697-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-040125

Quant Time: Apr 03 12:10:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66672	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129805	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51137	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65511	53.730	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.460%	
35) Dibromofluoromethane	5.379	113	47572	51.654	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.300%	
50) Toluene-d8	8.647	98	161884	50.360	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.720%	
62) 4-Bromofluorobenzene	11.079	95	59883	51.143	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.280%	
Target Compounds						Qvalue
16) Acetone	2.386	43	1055	2.107	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
Data File : VX045562.D
Acq On : 03 Apr 2025 11:45
Operator : JC/MD
Sample : Q1697-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-040125

Quant Time: Apr 03 12:10:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

