

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
 Data File : VX045564.D
 Acq On : 03 Apr 2025 12:31
 Operator : JC/MD
 Sample : Q1711-01 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18B-56-040225

Quant Time: Apr 04 01:20:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122051	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48003	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63046	54.791	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.580%			
35) Dibromofluoromethane	5.385	113	45176	52.169	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.340%			
50) Toluene-d8	8.647	98	155976	51.605	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.200%			
62) 4-Bromofluorobenzene	11.079	95	57430	52.164	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.320%			
Target Compounds						Qvalue
16) Acetone	2.386	43	642	1.359	ug/l	93
27) cis-1,2-Dichloroethene	4.495	96	3710	3.946	ug/l	96
44) Trichloroethene	7.135	130	1854	2.184	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
Data File : VX045564.D
Acq On : 03 Apr 2025 12:31
Operator : JC/MD
Sample : Q1711-01 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18B-56-040225

Quant Time: Apr 04 01:20:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

