

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
 Data File : VX045571.D
 Acq On : 03 Apr 2025 15:14
 Operator : JC/MD
 Sample : Q1711-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-040225

Quant Time: Apr 04 01:23:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129076	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121728	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49430	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65957	53.882	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.760%
35) Dibromofluoromethane	5.385	113	47665	52.048	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.100%
50) Toluene-d8	8.647	98	162893	50.960	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.920%
62) 4-Bromofluorobenzene	11.079	95	59432	51.044	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
Data File : VX045571.D
Acq On : 03 Apr 2025 15:14
Operator : JC/MD
Sample : Q1711-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-040225

Quant Time: Apr 04 01:23:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

