

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
 Data File : VX045575.D
 Acq On : 03 Apr 2025 16:47
 Operator : JC/MD
 Sample : Q1711-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18B-56-040225

Quant Time: Apr 04 01:25:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64116	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	121192	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112037	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	43792	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62586	53.378	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.760%	
35) Dibromofluoromethane	5.379	113	44291	51.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.020%	
50) Toluene-d8	8.647	98	153765	51.234	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.460%	
62) 4-Bromofluorobenzene	11.079	95	53724	49.144	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.280%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	4299	4.769	ug/l	96
12) 1,1-Dichloroethene	2.312	96	997	1.295	ug/l	88
16) Acetone	2.379	43	5722	11.885	ug/l	95
27) cis-1,2-Dichloroethene	4.483	96	81070	84.620	ug/l	90
29) Tetrahydrofuran	5.019	42	5635	12.332	ug/l	96
44) Trichloroethene	7.122	130	30963	36.740	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
Data File : VX045575.D
Acq On : 03 Apr 2025 16:47
Operator : JC/MD
Sample : Q1711-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18B-56-040225

Quant Time: Apr 04 01:25:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

