

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
 Data File : VX045578.D
 Acq On : 03 Apr 2025 17:56
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 04 01:28:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50893	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64538	52.698	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.400%
35) Dibromofluoromethane	5.379	113	47009	50.563	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.120%
50) Toluene-d8	8.647	98	165239	50.920	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.840%
62) 4-Bromofluorobenzene	11.079	95	60195	50.926	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
Data File : VX045578.D
Acq On : 03 Apr 2025 17:56
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 04 01:28:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

