

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
 Data File : VX045581.D
 Acq On : 03 Apr 2025 19:06
 Operator : JC/MD
 Sample : Q1676-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-5

Quant Time: Apr 04 01:29:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	65062	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128448	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48761	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65303	54.885	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.780%	
35) Dibromofluoromethane	5.385	113	46714	51.259	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.520%	
50) Toluene-d8	8.647	98	161223	50.684	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.360%	
62) 4-Bromofluorobenzene	11.079	95	59011	50.931	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.860%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
Data File : VX045581.D
Acq On : 03 Apr 2025 19:06
Operator : JC/MD
Sample : Q1676-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-5

Quant Time: Apr 04 01:29:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

