

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
 Data File : VX045582.D
 Acq On : 03 Apr 2025 19:29
 Operator : JC/MD
 Sample : Q1676-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Apr 04 01:29:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128935	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65479	53.803	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.600%	
35) Dibromofluoromethane	5.379	113	45954	50.234	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.460%	
50) Toluene-d8	8.647	98	161785	50.669	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.340%	
62) 4-Bromofluorobenzene	11.079	95	58361	50.179	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 100.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040325\
Data File : VX045582.D
Acq On : 03 Apr 2025 19:29
Operator : JC/MD
Sample : Q1676-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Apr 04 01:29:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

