

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040419\
 Data File : VX008631.D
 Acq On : 04 Apr 2019 13:24
 Operator : JC/SP
 Sample : K2263-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-040319

Quant Time: Apr 05 07:41:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	216612	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	366524	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	320488	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133281	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	153734	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
35) Dibromofluoromethane	5.50	113	112322	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
50) Toluene-d8	8.71	98	460025	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
62) 4-Bromofluorobenzene	11.13	95	156897	47.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.84%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040419\
Data File : VX008631.D
Acq On : 04 Apr 2019 13:24
Operator : JC/SP
Sample : K2263-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-040319

Quant Time: Apr 05 07:41:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

