

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040419\
 Data File : VX008635.D
 Acq On : 04 Apr 2019 14:57
 Operator : JC/SP
 Sample : K2263-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Apr 05 08:04:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	215851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	363484	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	310277	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130343	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	152338	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
35) Dibromofluoromethane	5.50	113	112188	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	8.71	98	455511	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	11.13	95	151670	46.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.44%	
Target Compounds						
16) Acetone	2.44	43	4732	2.610	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040419\
Data File : VX008635.D
Acq On : 04 Apr 2019 14:57
Operator : JC/SP
Sample : K2263-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Apr 05 08:04:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

