

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040419\
 Data File : VX008643.D
 Acq On : 04 Apr 2019 18:04
 Operator : JC/SP
 Sample : K2275-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S8-EB-2019-1

Quant Time: Apr 05 08:31:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	215763	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	361772	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	313445	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132240	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	149706	47.57	ug/l	0.00
Spiked Amount						
			Recovery	=	95.14%	
35) Dibromofluoromethane	5.50	113	109731	47.44	ug/l	0.00
Spiked Amount						
			Recovery	=	94.88%	
50) Toluene-d8	8.71	98	451351	49.01	ug/l	0.00
Spiked Amount						
			Recovery	=	98.02%	
62) 4-Bromofluorobenzene	11.13	95	153154	46.89	ug/l	0.00
Spiked Amount						
			Recovery	=	93.78%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	7076	3.904	ug/l	91
20) Methylene Chloride	2.85	84	93155	30.703	ug/l	96
25) 2-Butanone	4.68	43	15837	5.528	ug/l	93
29) Tetrahydrofuran	5.13	42	239833	126.130	ug/l	99
40) Benzene	6.15	78	10421	0.885	ug/l	97
52) Toluene	8.79	92	10819	1.558	ug/l	99
67) Ethyl Benzene	10.26	91	4091	0.316	ug/l	95
68) m/p-Xylenes	10.36	106	4225	0.907	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040419\
Data File : VX008643.D
Acq On : 04 Apr 2019 18:04
Operator : JC/SP
Sample : K2275-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S8-EB-2019-1

Quant Time: Apr 05 08:31:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

