

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040419\
 Data File : VX008646.D
 Acq On : 04 Apr 2019 19:14
 Operator : JC/SP
 Sample : K2275-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S8-0710

Quant Time: Apr 05 08:40:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	211301	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	351486	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	309611	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132055	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	145349	47.16	ug/l	0.00
Spiked Amount						
			Recovery	=	94.32%	
35) Dibromofluoromethane	5.50	113	106056	47.19	ug/l	0.00
Spiked Amount						
			Recovery	=	94.38%	
50) Toluene-d8	8.71	98	446765	49.93	ug/l	0.00
Spiked Amount						
			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.13	95	151910	47.87	ug/l	0.00
Spiked Amount						
			Recovery	=	95.74%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	3932	2.215	ug/l	94
52) Toluene	8.78	92	3079	0.456	ug/l	97
67) Ethyl Benzene	10.25	91	4893	0.382	ug/l	96
68) m/p-Xylenes	10.36	106	6055	1.316	ug/l	98
95) Naphthalene	13.83	128	4102	0.377	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040419\
Data File : VX008646.D
Acq On : 04 Apr 2019 19:14
Operator : JC/SP
Sample : K2275-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S8-0710

Quant Time: Apr 05 08:40:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

