

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040419\
 Data File : VX008657.D
 Acq On : 05 Apr 2019 00:16
 Operator : JC/SP
 Sample : K2274-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006RW116-190402

Quant Time: Apr 05 11:02:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	211666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	353347	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	301561	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	127011	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	145313	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
35) Dibromofluoromethane	5.49	113	107420	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
50) Toluene-d8	8.71	98	436983	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
62) 4-Bromofluorobenzene	11.13	95	148164	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1250	0.380	ug/l	94
9) 1,1,2-Trichlorotrifluoroet	2.39	101	3968	1.693	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	101914	33.489	ug/l	89
30) Chloroform	5.21	83	1440	0.302	ug/l	95
44) Trichloroethene	7.21	130	22253	8.303	ug/l	100
64) Tetrachloroethene	9.34	164	129244	57.847	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040419\
Data File : VX008657.D
Acq On : 05 Apr 2019 00:16
Operator : JC/SP
Sample : K2274-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006RW116-190402

Quant Time: Apr 05 11:02:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

