

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040419\
 Data File : VX008660.D
 Acq On : 05 Apr 2019 01:26
 Operator : JC/SP
 Sample : K2274-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW079-190403

Quant Time: Apr 05 09:14:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	219007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	364484	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	316919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133924	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	151611	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
35) Dibromofluoromethane	5.49	113	111186	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
50) Toluene-d8	8.71	98	453869	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
62) 4-Bromofluorobenzene	11.13	95	154432	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	49520	15.727	ug/l	90
44) Trichloroethene	7.21	130	22921	8.291	ug/l	93
64) Tetrachloroethene	9.34	164	11567	4.926	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040419\
Data File : VX008660.D
Acq On : 05 Apr 2019 01:26
Operator : JC/SP
Sample : K2274-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW079-190403

Quant Time: Apr 05 09:14:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

