

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040419\
 Data File : VX008662.D
 Acq On : 05 Apr 2019 02:12
 Operator : JC/SP
 Sample : K2274-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW012-190402

Quant Time: Apr 05 09:18:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	221325	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	363816	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	313056	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133853	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	151309	46.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.74%	
35) Dibromofluoromethane	5.50	113	110830	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
50) Toluene-d8	8.71	98	463181	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	11.13	95	155982	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	47106	13.678	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	29701	9.334	ug/l	89
39) Methylcyclohexane	7.45	83	5492	1.118	ug/l #	44
44) Trichloroethene	7.22	130	3713	1.346	ug/l	95
64) Tetrachloroethene	9.34	164	8513	3.670	ug/l	97
83) tert-Butylbenzene	11.77	119	3174	0.354	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040419\
Data File : VX008662.D
Acq On : 05 Apr 2019 02:12
Operator : JC/SP
Sample : K2274-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW012-190402

Quant Time: Apr 05 09:18:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

